

Message

From: Mark Pandori [/O=THERANOS ORGANIZATION/OU=EXCHANGE ADMINISTRATIVE GROUP (FYDIBOHF23SPDLT)/CN=RECIPIENTS/CN=MARK PANDORI16D]
Sent: 2/10/2014 10:27:26 PM
To: Adam Rosendorff [/o=theranos organization/ou=exchange administrative group (fydibohf23spdlT)/cn=recipients/cn=adam rosendorfd92]
Subject: alert: FW: QC for TSH Failed After Re-run

Alert.

The TSH is for a potential investor.

Mark

From: Erika Cheung
Sent: Monday, February 10, 2014 2:19 PM
To: Adam Rosendorff; Jamie Liu
Cc: Mark Pandori
Subject: QC for TSH Failed After Re-run

Hi Adam,

I re-ran the level 1 QC for TSH, it failed again. The data is in the tab titled QC_2013.1.23-current under the row with today's date on it.

If we need to recalibrate, I'm just going to use the TSH CAPSYS cartridges. I should be finished running the daily QC information and getting the calibration finished by tomorrow or Wednesday. If this is the case, expect a delay on processing the patient sample until then.

Thanks,

Erika