

To: Elizabeth Holmes[eholmes@theranos.com]
Cc: Sunny Balwani[sbalwani@theranos.com]; Christian Holmes[cholmes@theranos.com]
From: Daniel Edlin
Sent: Sat 6/8/2013 10:44:16 PM
Importance: Normal
Subject: Updated Presentation Deck Template
Received: Sat 6/8/2013 10:44:25 PM
Theranos Presentation Template June2013.ppt

Hi Elizabeth,

Please find attached the updated presentation deck template including only those slides that you've focused on/spoken to in recent presentations. Let us know if you'd like to see any slides removed or added back in.

Saved here: J:\Product Management - Commercial\Presentations\Template\Theranos_Presentation_Template_June2013.ppt

Thanks,

Dan

File Produced in Native Format



INSERT PARTNER/AUDIENCE LOGO

INSERT DATE

Contents

Background on Theranos

Transforming the Clinical Laboratory

Cost Savings & Hospital System Partnership

Clinical Deep-Dive

the lab test, reinvented.



same tests.
smaller sample.



Theranos, Inc.

Theranos is a Silicon Valley-based healthcare technology company founded in 2003.

Theranos' proprietary, patented technology runs comprehensive blood tests from a finger-stick and tests from micro-samples of other matrices in real-time outside of traditional lab settings and generates significantly higher integrity data than currently possible.

Our current and past clients include 10 of the top 15 major pharmaceutical companies, midsized bio-pharmas, prominent research institutions, healthcare payors, and U.S. and foreign government health and military organizations.

Contents

Background on Theranos

Transforming the Clinical Laboratory

Cost Savings & Hospital System Partnership

Clinical Deep-Dive

Transforming the Clinical Lab

Check-In


Prep & Perform
Finger Stick

Collect
Biometrics

Close-Out

Lab Today

Theranos



Finger-stick tests **reduce volumes of blood draws by 99%**

Major impact on **patient experience**: pediatrics, geriatrics, oncology, etc.

Excerpts from Theranos' Test Menu

*105 tests shown, another 20+ pages show all available tests with Theranos

<u>Bacteria</u>	<u>Complete Blood Count w Diff</u>		<u>Complete Metabolic Panel</u>	<u>Cardiovascular Panel</u>
Streptococcus pneumoniae (penic R(24%),S)	White blood cell count		HGB A1c	Creatinine
Mycoplasma pneumoniae	Red blood cell count		Glucose	Kinase
Chlamydia pneumoniae	Hemoglobin		Calcium	Troponin-I
Bordetella pertussis	Hematocrit		Albumin	Troponin-t
Haemophilus influenzae (ampic R,S)	Mean corpuscular volume		Total Protein	CRP- High-Sensitivity & LS
Moraxella catarrhalis	Mean corpuscular hemoglobin		Sodium	Homocysteine
Staphylococcus aureus (MR (30%), RS)	Mean corpuscular hemoglobin concentration		Potassium	
Streptococcus pyogenes (A)	Platelet count		CO2	
Streptococcus agalactiae (B)	Mean platelet volume		Chloride	<u>Lipid Profile & Glucose Panel</u>
Pseudomonas spp (aeruginosa)		<u>Renal Panel</u>	BUN	Cholesterol
Haemophilus parainfluenzae			Creatinine	HDL
Enterobacteriaceae spp	Albumin	ALT	ALP	LDL
Legionella spp	BUN	Alkaline Phosphatase	AST	LDL/HDL Ratio
gram-negative bacteria	Calcium	AST	AST	Triglycerides
Escherichia coli	CO2	Ferritin	Bilirubin	VLDL
	Chloride	GGT	Magnesium	
<u>Viral</u>	Glucose	Iron		<u>STDs & Drugs of Abuse</u>
H5N1, H1N1	Phosphorous	Lactate Dehydrogenase	Ipecac	Chylimd Trach, Dna, Amp Probe
H3N2, Infl. B	Potassium	Microalbumin	Lsd,	N.Gonorrhoeae, Dna, Amp Prob
Rhino Virus	Sodium	Total Protein	Lsd-25,	Hpv, Dna, Amp Probe
Adenovirus	Creatinine	Albumin	Lysergide,	Acid
RSV	eGFR	Globulin	Nalbuphine	Butorphanol
parainfluenza virus (1,2,3,4)		Bilirubin Direct	Nubain(R)	D-Lysergicacid Diethylamide,
Coronaviruses	TSH	Bilirubin Total	Rohipnolâ®	Dolophine,
human metapneumovirus (HMPV)	T-3		Stadolâ®	Flunitrazepam
	T-4		Ethyl Glucuronide,	Heroin,

Theranos is certified as a High Complexity CLIA Laboratory

Waived

Simple, accurate tests without routine oversight

Moderate

Most tests fall in this category; automated testing where the lab must meet standards and surveyed biennially

PPM

Provider performed microscopy; the lab must meet quality standards; no routine oversight

High Complexity

Requires the highest level of training, technique and result interpretation; most stringent standards; labs are surveyed routinely and randomly

State of California

Department of Public Health

CLINICAL LABORATORY LICENSE

In accordance with the provisions of Chapter 3, Division 2 of the Business and Professions Code,
the persons named below are hereby issued a license authorizing operation of a clinical laboratory
at the indicated address or other site(s) on file with the department.

THERANOS, INC.
1001 EAST MEADOW CIRCLE
PALO ALTO CA 94303-4231

OWNER(S):

THERANOS, INC.

DIRECTOR(S):


ARNOLD B. GELB MD
SPENCER HIRAKI PHD

Lab ID Number: CLF 00341367

Effective Date: June 15, 2012

Valid Until: June 14, 2013

CLIA Number: 05D2025714


Beatrice O'Keefe
Beatrice R. O'Keefe, Division Chief
Laboratory Field Services

Validation of Theranos

Theranos has been comprehensively validated over the course of the last seven years by ten of the fifteen largest pharmaceutical companies, with hundreds of thousands of assays processed.

After running clinical trials with Theranos instead of the central laboratory, GlaxoSmithKline's Lab Director concluded that **"Theranos' lab infrastructure eliminates the need for a lab."**

Theranos' lab infrastructure is validated under   ICH, and  World Health Organization guidelines.

Excerpts from Johns Hopkins due diligence and technology validation:

- “The technology is novel and sound. It can accurately run a wide range of routine and special assays.”
- “**No major weaknesses were identified.**”



Theranos Information Systems

Theranos Information Systems facilitate **real-time eligibility, authorization, authentication, information transmission, and billing**

All data is transmitted to physicians through a **secure customized portal, secure fax, and/or integration with EMR/LIS systems**

Theranos Patient Service Centers



Theranos' Footprint at retail:

Theranos patient service centers are located within a smaller radius from the patient than currently available

Theranos has more patient service centers than any lab provider in CA

Convenience is offered at an unprecedented value

	1 mile	3 miles	5 miles
theranos	> 95%		
 Quest Diagnostics	7%	35%	56%
 LabCorp <small>Laboratory Corporation of America</small>	9%	45%	69%

Theranos Sites



theranos









Contents

Background on Theranos


Transforming the Clinical Laboratory

Cost Savings & Hospital System Partnership

Clinical Deep-Dive

Cost Savings: NY Medicare/Medicaid


Estimated Direct out-of-pocket Cost Savings
for New York **Medicaid**



10-year aggregate savings of
\$22.3 billion

Source: CMS.gov, KFF.org and Theranos estimates

Estimated Direct out-of-pocket Cost Savings
for New York **Medicare**



10-year aggregate savings of
\$6.3 billion

Contents

Background on Theranos


Transforming the Clinical Laboratory

Cost Savings & Hospital System Partnership

Clinical Deep-Dive

Theranos Data Infrastructure


Standards-based import tools allow for automatic importation, integration, and standardization of data from any and all clinical databases.




Predictive Insight: Disease Progression

Robust studies have shown that more frequent sampling on a low variability platform allows characterization of trends that cannot be seen when patients come into the clinic for blood draws less frequently and run in traditional labs.

Variability in M30 and M65 Pre-dose Levels
(5-7 day gap between 2 pre-dose samples)



Time series: chemo-sensitive solid tumor and M30 M65 trends




Algorithms used to reduce false positive results through large scale pattern analysis

Case study

95% predictive power

The model takes into account complex physiological system pathways

Model recognizes signatures of BMD changes measurable ~6 months prior to physical changes in BMD to better assess need for intervention



Population Centric and Disease Specific Applications Modeling & Simulation of H1N1: District Federal

Based on large set of clinical records from patients with influenza-like illness (ILI) in Mexico from 2009:

Utilized the Theranos influenza model to simulate the H1N1 outbreak in an urban, dense population


Evaluated the impact and cost-effectiveness of enhanced testing and surveillance using THS

Assessed vaccination strategies to determine the impact on the number of cases and deaths

Explored the impact of different surveillance strategies under different household secondary attack rates


Theranos Health Threat Control Center for IMMS

Predictions for H1N1 outbreak and recommended interventions



Attack rate: number of suspected cases per 1 million people
Mexico City shows later but sustained high attack rate

Model Predicted Spread of Outbreak w/ >99% Accuracy



immunochemistry

ANA Antibody

Sample ID	Theranos Ab Index	Immco Diagnostics ANA value	USBIOL Ab index	Innova Ab index	IBL Ab Index	ProMeddx Nephelometry ratio
A1	0.77	39.1	0.53	12.2	0.82	1:160
A2	0.73	23.6	0.73	10.4	0.49	1:320
A3	1.78	36.6	0.90	31.2	0.76	1:320
A4	2.36	34.3	1.07	21.5	0.72	1:160
A5	0.62	12.8	0.52	13.3	0.27	1:160
A6	1.14	39.9	1.26	13.3	0.83	1:160
A7	1.64	27.0	0.58	8.4	0.56	1:320
A8	1.89	22.5	0.69	8.4	0.47	1:320
A9	2.18	143.8	2.41	39.1	3.00	1:1280
A10	6.14	32.8	2.13	21.2	0.69	1:1280
A11	1.68	74.1	2.50	74.5	1.55	1:1280
A12	19.53	157.3	2.78	106.2	3.29	1:1280
A13	1.51	32.8	2.75	20.5	0.69	1:160
A14	3.16	31.8	1.85	19.2	0.66	1:160
A15	1.92	24.4	1.01	13.8	0.51	1:320
A16	1.85	58.5	5.36	39.4	1.22	1:1280
A17	5.57	153.8	3.81	99.8	3.21	1:1280
A18	1.45	33.4	1.03	25.1	0.70	1:640
A19	7.98	99.6	2.83	81.5	2.08	1:640
A20	1.87	51.4	0.86	21.3	1.07	1:640
A21	1.06	50.0	0.83	11.9	1.04	1:640
A22	1.38	31.7	1.16	23.2	0.66	1:640
A23	0.95	53.8	1.00	14.0	1.12	1:320
A24	1.57	38.1	1.41	13.7	0.80	1:1280
A25	1.65	55.6	2.55	17.8	1.16	1:640

Sample ID	Theranos Ab Index	Immco Diagnostics ANA value	USBIOL Ab index	Innova Ab index	IBL Ab Index	ProMeddx Nephelometry ratio
A26	0.87	43.2				1:640
A27	1.69	28.2				1:640
A28	1.85	37.9				1:640
A29	0.92	23.4				1:160
A30	3.19	87.4				1:160
A32	1.68	48.9				1:160
A33	1.23	45.3				1:1280
A34	2.10	50.0				1:1280
A35	0.85	31.0				1:320
A36	0.54	40.0				1:160

Results reported as:
 Negative (green), Equivocal (Yellow), or
 Positive (Red) on the basis of Antibody
 index for each assay.
 Immco kit is an FDA approved ELISA

Ab Index > 1.1
Ab Index > 0.9, < 1.1
Ab Index < 0.9

Anti - RNP

Sample ID	Human Test Samples			ANTIBODY INDEX			
	Matrix	Species	Strain	Theranos	INOVA	Corgenix	IBL
CLN1	Serum	Normal	N/A	0.01	4	3.28	0.19
CLN2	Serum	Normal	N/A	0.01	4	1.64	0.08
CLN3	Serum	Normal	N/A	0.01	6	3.31	0.27
CLN4	Serum	Normal	N/A	0.01	5	2.77	0.08
CLN5	Serum	Normal	N/A	0.01	4	2.89	0.06
CLN6	Serum	Normal	N/A	0.01	7	2.58	0.16
CLN7	Serum	Normal	N/A	0.01	6	3.33	0.19
CLN8	Serum	Normal	N/A	0.01	5	5.67	0.19
CLN9	Serum	Normal	N/A	0.01	4	1.69	0.12
CLN10	Serum	Normal	N/A	0.01	5	3.32	0.12
SL04	Serum	Autoimmune	Lupus	0.07	38	76.86	3.08
SL07	Serum	Autoimmune	Lupus	3.46	150	226.67	8.08
SL08	Serum	Autoimmune	Lupus	0.03	52	35.77	2.26
SL09	Serum	Autoimmune	Lupus	3.10	148	227.45	8.19
CSLE2	Serum	Autoimmune	Lupus	0.02	98	18.02	0.03
CSLE4	Serum	Autoimmune	Lupus	3.32	142	155.73	4.57
CSLE5	Serum	Autoimmune	Lupus	0.45	90	17.69	0.44
CSLE7	Serum	Autoimmune	Lupus	3.21	146	124.13	3.17
CSLE8	Serum	Autoimmune	Lupus	0.39	126	106.37	1.38
CSLE10	Serum	Autoimmune	Lupus	0.96	142	88.19	2.24
CSLE11	Serum	Autoimmune	Lupus	1.15	140	154.59	4.58
CSLE13	Serum	Autoimmune	Lupus	0.02	88	5.56	0.24
CSLE14	Serum	Autoimmune	Lupus	0.13	33	63.84	1.61
CSLE15	Serum	Autoimmune	Lupus	2.85	147	164.33	2.10
SCL02	Serum	Autoimmune	Scleroderma	0.14	32	3.73	0.11
SCL05	Serum	Autoimmune	Scleroderma	0.05	13	23.55	0.91
SCL07	Serum	Autoimmune	Scleroderma	0.15	71	N/A	0.67
SCL11	Serum	Autoimmune	Scleroderma	0.12	35	59.46	1.81
SCL14	Serum	Autoimmune	Scleroderma	6.02	145	163.45	2.47
SCL35	Serum	Autoimmune	Scleroderma	0.01	N/A	21.16	N/A

Results reported as: Negative (green), Equivocal (Yellow), or Positive (Red) on the basis of Antibody index for each assay.

Anti - Sm

Samples	Inter-Cartridge		Theranos Ab	IMMCO Result
	Mean	CV%		
SL4	9842	15	1.07	14
SL7	17081	11	1.85	30
SL9	9210	8	1.00	26
Sjog7	5095	16	0.55	12
Sjog9	3153	2	0.34	26
Sjog10	3250	3	0.35	18
Scleroderma 1	5774	18	0.6	13
Scleroderma 2	4931	41	0.5	8
Scleroderma 3	7841	16	0.9	6
Scleroderma 4	6169	4	0.7	5
Scleroderma 5	18168	22	2.0	22
Scleroderma 6	5064	62	0.5	7
Scleroderma 7	7702	9	0.8	17
Scleroderma 8	8490	4	0.9	7
Scleroderma 9	5156	19	0.6	7
Scleroderma 10	8903	3	1.0	7
Scleroderma 11	19170	10	2.1	24
	22220			
Scleroderma 14	2	6 ³¹	24.1	29


Results reported as: Negative (green),
Equivocal (Yellow), or Positive (Red) on
the basis of Antibody index for each
assay.

Anti - SSB


Samples	Inter-Cartridge		Theranos	INOVA ELISA	IBL International
	Mean	CV%			
A7	7599	19	0.74	2.8	0.44
A13	4448	27	0.44	3.0	0.59
A14	5730	17	0.56	4.8	0.37
A15	5730	20	0.56	2.7	0.32
A17	7288	17	0.71	2.7	0.55
pooled Ana positives	5751	4	0.56	2.7	0.55
Sj01	565	3	0.06	2.8	0.24
Sj02	806	6	0.08	3.0	0.16
Sj03	120175	12	11.77	62.2	2.62
Sj04	626	17	0.06	2.8	0.17
Sj05	204698	16	20.06	96.2	4.67
Sj06	2398	15	0.23	2.8	0.20
Sj07	8161	12	0.80	10.9	0.24
Sj08	4873	2	0.48	17.6	0.24
Sj09	26981	24	2.64	36.3	0.26
Sj10	2500	6	0.24	15.1	0.53

Results reported as: Negative (green), Equivocal (Yellow), or Positive (Red) on the basis of Antibody index for each assay.


Acid Phosphatase – Serum & Plasma



Ascorbic Acid in Plasma




Trypanosoma cruzi (Chagas) antibody




theranos

CRP Test Correlation to Reference Methods Over 10,000-fold Range



Cotinine (Serum)



Cotinine (Urine)

Theranos Result	FDA approved	
	Cot 1 Cotinine	NicChek I
	ng/mL	Rapid test
OORH	Pos	LowPos
OORH	Pos	LowPos
OORL	Neg	Neg
OORH	Pos	LowPos
765	Pos	LowPos
OORH	Pos	LowPos
OORH	Pos	LowPos
OORL	Neg	Neg
894	Pos	LowPos
1793	Pos	LowPos
OORL	Neg	Neg
178	Pos	LowPos
347	Pos	LowPos
OORH	Pos	High pos
OORL	Neg	Neg
OORH	Pos	High pos
OORH	Pos	High pos

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

CMV IgG

	Theranos cutoff		Trinity (C of A)	DiaSorin (CLIA)
	5*stdev	10*stdev		
Zepto MT Panel #1	0.51	0.31	0.3	<<0.20
Zepto MT Panel #2	39.02	23.99	2.4	3.1
Zepto MT Panel #3	56.76	34.89	2.7	6.1
Zepto MT Panel #4	45.48	27.95	2.4	3.3
Zepto MT Panel #5	17.69	10.87	2.8	8.4
Zepto MT Panel #6	45.25	27.82	2.8	5.9
Zepto MT Panel #7	65.68	40.37	2.8	>>10.0
Zepto MT Panel #8	62.07	38.16	2.8	>>10.0
Zepto MT Panel #9	63.62	39.11	2.8	>>10.0
Zepto MT Panel #10	27.03	16.62	2.8	4.8
Zepto MT Panel #11	0.74	0.45	0.2	<<0.20
Zepto MT Panel #12	29.42	18.09	3.2	6.8
Zepto MT Panel #13	10.27	6.31	1.9	3.8
Zepto MT Panel #14	86.07	52.91	2.8	>>10.0
Zepto MT Panel #15	30.47	18.73	2.8	6.1
Zepto MT Panel #16	0.73	0.45	0.3	<<0.20
Zepto MT Panel #17	1.18	0.72	0.1	<<0.20
Zepto MT Panel #18	76.77	47.19	2.7	7.8
Zepto MT Panel #19	1.12	0.69	0.2	<<0.20
Zepto MT Panel #20	9.82	6.04	1.9	4.4
Zepto MT Panel #21	1.28	0.78	0.1	<<0.20
Zepto MT Panel #22	33.62	20.67	3.3	>>10.0
Zepto MT Panel #23	11.86	7.29	2.1	2.5
Zepto MT Panel #24	16.62	10.22	2.1	1.8

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

Dengue IgG – Clinical correlation

Result	Calbiotech	Genway	InBios	THERANOS
	Qualitative interpretation:		ISR	ISR
Negative	<0.9	<0.9	< and = to 1.65	< and = to 1.65
Equivocal	0.9 - 1.1	0.9 - 1.1	1.65 - 2.84	1.65 - 2.84
Positive	>1.1	>1.1	> and = to 2.84	> and = to 2.84

NORMALS		(FDA approved)		
Sample	Calbiotech	Genway	InBios	THERANOS
Sample	Results	Results	RA/NCA Ratio	RA/NCA Ratio
Nor01	0.180	0.183	1.279	2.257
Nor02	0.135	0.142	1.127	0.786
Nor03	0.057	0.071	1.027	0.770
Nor04	0.047	0.062	0.949	0.465
Nor05	0.048	0.048	1.060	0.319
Nor06	0.049	0.047	0.956	0.787
Nor07	0.128	0.129	0.934	0.763
Nor08	0.142	0.147	0.874	0.401
Nor09	0.039	0.038	1.304	0.353
Nor10	0.126	0.144	0.940	0.407
Nor11	0.032	0.034	1.030	0.506
Nor12	0.301	0.377	0.848	0.278
Nor13	0.048	0.046	1.033	0.256
Nor14	0.067	0.072	0.864	0.256
Nor15	0.161	0.159	0.899	0.321
Nor16	0.332	0.457	0.862	0.320
Nor17	0.116	0.173	1.253	0.336
Nor18	0.334	0.416	0.883	0.494
Nor19	0.048	0.047	0.881	0.285
Nor20	0.113	0.126	1.018	0.698

Dengue IgG – Clinical correlation (contd.)

RF/Hama			(FDA approved)	
	Calbiotech	Genway	InBios	THERANOS
Sample	Results	Results	RA/NCA Ratio	RA/NCA Ratio
Rf (#1)	0.17	0.16	1.33	1.26
Rf (#2)	0.14	0.14	1.08	1.71
Rf (#3)	0.13	0.15	1.23	1.92
Rf (#4)	0.17	0.05	0.94	2.23
Rf (#5)	0.13	0.05	1.33	0.83
Rf (#6)	0.05	0.18	0.83	0.81
Hama (#1)	0.14	0.14	0.78	0.73
Hama (#2)	0.14	0.16	0.84	1.12
Hama (#3)	0.04	0.16	1.12	0.88
Hama (#4)	0.13	0.15	1.01	0.67
Hama (#5)	0.03	0.04	1.02	1.46
Hama (#6)	0.28	0.37	0.94	1.50

Results reported as: Negative (green), Equivocal (Yellow), Positive (Red), on the basis of Antibody index for each assay.

Dengue IgG – Clinical correlation (contd.)

			Calbiotech Dengue Virus IgG ELISA	EUROIMMUN Anti-Dengue Virus ELISA (IgG)	Focus Dengue Virus IgG DiSelect	Panbio Dengue IgG Capture ELISA	SD Dengue IgG Capture ELISA	THERANOS
Sample	Country of Origin	Seracare Panel	Results	Results	Results	Results	Results	R:M:CARatio
Clinical (#1)	Colombia	PVD201-01	15	49	62	24	>8.9	80.7
Clinical (#2)	Honduras	PVD201-02	18	53	5.7	12	16	81.5
Clinical (#3)	Honduras	PVD201-03	22	48	7.1	13	11	91.0
Clinical (#4)	Honduras	PVD201-04	1.7	20	3.6	0.5	0.5	40.4
Clinical (#5)	Honduras	PVD201-05	21	5.1	7.0	4.1	1.9	110.4
Clinical (#6)	Honduras	PVD201-06	23	44	63	4.5	3.4	86.4
Clinical (#7)	Colombia	PVD201-07	13	4.7	7.0	7.0	8.8	117.3
Clinical (#8)	Honduras	PVD201-08	23	43	65	26	12	40.5
Clinical (#9)	Honduras	PVD201-09	21	44	5.7	0.7	2.1	35.5
Clinical (#10)	Ecuador	PVD201-10	13	3.8	2.2	0.4	2.5	42.2
Clinical (#11)	Honduras	PVD201-11	2.1	4.2	4.8	0.5	1	84.1
Clinical (#12)	Honduras	PVD201-12	21	3.0	5.3	0.9	0.7	49.3
Clinical (#13)	Honduras	PVD201-13	20	47	58	13	2.1	151.4
Clinical (#14)	USA	PVD201-14	0.4	0.1	0.0	0.0	0.1	0.5
Clinical (#15)	Honduras	PVD201-15	24	3.7	5.2	13	0.8	82.4
Clinical (#16)	Ecuador	PVD201-16	14	41	25	0.7	2.4	53.9
Clinical (#17)	Colombia	PVD201-17	15	5.2	5.9	23	>8.9	116.7
Clinical (#18)	Honduras	PVD201-18	1.9	3.1	5.4	23	0.7	58.4
Clinical (#19)	Honduras	PVD201-19	22	3.8	5.8	0.6	0.5	51.1
Clinical (#20)	Ecuador	PVD201-20	1.2	4.4	2.2	0.7	3	57.0
Clinical (#21)	Ecuador	PVD201-21	17	54	68	43	>8.9	45.5

Positive samples from Seracare

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

Dengue IgG – Clinical correlation (contd.)

Sample #	Calbiotech EIA	Euroimmune EIA	Focus EIA	Panbiotech EIA	Panbiotech Indirect EIA	Standard Diagnostics EIA	THERANOS
9253423	1.663	5.614	6.938	4.654	3.947	8.929	152.6
9253421	1.204	4.541	2.28	0.886	3.174	3.036	89.5
9253422	1.265	4.106	2.333	0.516	0.007	2.506	41.0
9253425	1.372	4.211	2.484	0.977	3.209	2.423	77.0
9254165	1.337	5.211	7.071	7.487	3.947	8.771	84.9
9254166	1.463	4.988	6.298	2.922	3.947	8.929	109.8
9254167	1.49	5.496	6.036	2.775	3.947	8.929	101.5
9240601	2.609	4.992	6.547	2.355			73.3
9242868					2.703		22.9
9254539					3.525		80.2
9254540					3.716		90.6
9254542					3.947		119.6
9254543					3.947		132.2
9256502	2.201						98.0
9256614	3.626						110.1

Results reported as: Negative (green) or Positive (Red) on the basis of Antibody index for each assay.

Dengue IgM Clinical correlation

SeraCare DENV Sample	Theranos	Focus DxSelect	InBios DENV	Panbio
PVD201-01	POS	POS	EQUIVOCAL	POS
PVD201-02	NEG	NEG	NEG	NEG
PVD201-03	NEG	NEG	NEG	EQUIVOCAL
PVD201-04	NEG	NEG	NEG	NEG
PVD201-05	NEG	NEG	NEG	NEG
PVD201-06	NEG	NEG	NEG	NEG
PVD201-07	POS	POS	POS	POS
PVD201-08	NEG	NEG	NEG	EQUIVOCAL
PVD201-09	NEG	NEG	NEG	NEG
PVD201-10	POS	POS	EQUIVOCAL	POS
PVD201-11	NEG	NEG	NEG	NEG
PVD201-12	NEG	NEG	NEG	NEG
PVD201-13	NEG	NEG	NEG	NEG
PVD201-14	NEG	NEG	NEG	NEG
PVD201-15	NEG	NEG	NEG	NEG
PVD201-16	EQUIVOCAL	NEG	NEG	POS
PVD201-17	POS	POS	EQUIVOCAL	POS
PVD201-18	NEG	NEG	NEG	NEG
PVD201-19	NEG	NEG	NEG	NEG
PVD201-20	POS	POS	EQUIVOCAL	POS
PVD201-21	POS	POS	EQUIVOCAL	POS
RESULTS	Theranos	Focus DxSelect	InBios DENV	Panbio
	<0.90 Negative	<1.00 Negative	<1.65 Negative	<0.9 Negative
	≤1.10 - ≥0.90 Equivocal		1.65-2.84 Equivocal	0.9 - 1.1 Equivocal
	>1.10 Positive	>1.00 Positive	>2.84 Positive	>1.1 Positive

ENA – Jo-1 antibody

	Innova	Immco	IBL Int.	Theranos
Sample	Units	Result (EU/mL)	Ratio	Ab Index
1	0	6	0.1	0.6
2	1	7	0.1	0.7
3	32	126	4.4	5.7
4	127	177	6.6	18.1
5	111	169	6.6	9.6
6	5	8	0.1	0.7
7	19	4	0.1	0.0
8	3	10	0.4	0.6
9	8	13	0.3	0.7
10	1	14	0.2	0.7
11	2	12	0.3	0.0
12	1	7	0.1	0.0
13	0	5	0.1	0.0
14	1	3	0.1	0.0
15	1	20	0.2	0.3
16	0	8	0.1	0.0
17	165	175	6.2	16.6
18	1	8	0.2	0.0
19	50	138	4.4	8.1
20	9	10	0.2	0.0
21	107	175	6.3	8.9
22	8	8	0.1	0.0
23	3	10	0.4	0.0
24	5	14	0.3	0.5
25	23	2	0.1	0.4
26	107	201	7.9	2.5
27	69	158	6.5	2.3
28	110	203	8.1	2.6
29	108	208	8.2	2.7
30	109	201	8.2	3.0

Ab index >1.1	Positive
Ab index >0.9<1.1	Equivocal
Ab index < 0.9	Negative


Epstein Barr Virus-Early Antigen (EBV-EA) IgG

Sample ID	Theranos Index	Liaison Result
PE19	1.00	Negative
B3	2.07	Negative
M1	1.18	Positive
M2	5.43	Positive
M3	1.30	Positive
M4	3.70	Positive
M5	0.44	Negative
M6	3.27	Positive
M7	0.50	Negative
M8	1.19	Positive
M9	0.37	Positive
M10	9.81	Positive
M11	33.26	Positive
M12	19.73	Positive
M13	23.50	Positive
M14	30.30	Positive
M15	0.71	Negative


Sample ID	Theranos Index	Liaison Result
M16	15.13	Positive
M17	4.75	Positive
M18	0.68	Positive
M19	0.56	Negative
M20	0.70	Negative
M21	2.07	Positive
C12	16.97	Positive
C13	16.44	Positive
C14	6.98	Positive
C15	25.11	Positive
C16	24.31	Positive
C17	6.94	Positive
C18	23.52	Positive
C19	17.07	Positive
C20	6.07	Positive
C22	29.08	Positive

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

Ferritin




Follicle stimulating Hormone (FSH)



Anti-HAV IgM

	Theranos		Siemens Advia Centaur		Abbott Architect		DiaSorin-ETI-HA-IGMK Plus	
Sample	S/CO	Result	S/CO	Result	S/CO	Result	S/CO	Result
1	0.4	Negative	0.3	Negative	0.8	Negative	0.1	Negative
2	0.3	Negative	0.2	Negative	0.4	Negative	0.1	Negative
3	0.3	Negative	0.2	Negative	0.5	Negative	0.0	Negative
4	0.2	Negative	0.2	Negative	0.3	Negative	0.1	Negative
5	6.7	Positive	> 7.0	Positive	2.8	Positive	6.1	Positive
6	4.5	Positive	4.1	Positive	1.1	Positive	3.8	Positive
7	4.4	Positive	4.0	Positive	1	Positive	3.5	Positive
8	0.4	Negative	0.3	Negative	0.7	Negative	0.1	Negative
9	0.2	Negative	0.2	Negative	0.4	Negative	0.1	Negative
10	0.3	Negative	0.4	Negative	0.4	Negative	0.1	Negative
11	8.0	Positive	> 7.0	Positive	3.6	Positive	6.5	Positive
12	0.2	Negative	0.2	Negative	0.4	Negative	0.0	Negative
13	0.8	Negative	0.3	Negative	0.7	Negative	0.1	Negative
14	2.6	Positive	2.5	Positive	0.6	Negative	1.3	Positive
15	5.8	Positive	5.9	Positive	1.5	Positive	5.2	Positive
16	0.5	Negative	0.4	Negative	0.4	Negative	0.1	Negative
17	9.6	Positive	6.5	Positive	2.1	Positive	5.3	Positive
18	0.2	Negative	0.2	Negative	0.5	Negative	0.0	Negative
19	6.8	Positive	3.4	Positive	1.7	Positive	3.0	Positive
20	1.8	Positive	0.2	Negative ⁴⁹	0.6	Negative	0.1	Negative
21	0.5	Negative	0.3	Negative ⁴⁹	0.4	Negative	0.1	Negative

HbA1c



DDL: Diabetes Diagnostic Lab – NGSP standardization program

Anti-HBc IgM

Seracare anti-HBc IgM Mixed Titer Performance Panel	Theranos Anti-HBc IgM (R > 2)	DPC Immulite Anti-HBc IgM (R >10 u/mL)	Abbott Architect anti-HBc IgM (R >1)	Diasorin ETI-CORE_IgMK (R>1)
	S/Co	PEI U/mL	S/Co	S/Co
Seracare 1	0.2	<2.0	0.0	0.1
Seracare 2	9.8	>100	12.9	10.9
Seracare 3	12.3	>100	22.6	11.9
Seracare 4	4.0	15.2	0.9	2.0
Seracare 5	11.1	>100	6.0	9.2
Seracare 6	7.9	88.2	1.8	3.6
Seracare 7	12.3	>100	20.3	12.5
Seracare 8	8.3	98.7	2.1	3.9
Seracare 9	12.2	>100	19.0	11.8
Seracare 10	8.8	>100	3.2	5.9
Seracare 11	10.1	>100	6.6	11.3
Seracare 12	1.8	8.8	0.1	0.6
Seracare 13	6.7	>100	2.8	6.2
Seracare 14	11.8	>100	18.2	11.8
Seracare 15	13.0	>100	6.7	10.9
Seracare 16	9.1	>100	4.6	6.1
Seracare 17	1.4	7.7	0.3	0.9
Seracare 18	4.8	37.9	1.5	2.6
Seracare 19	12.1	>100	24.6	12.7
Seracare 20	6.5	57.7	1.4	0.7
Seracare 21	6.7	59.6	1.7	3.3
Seracare 22	5.6	32.0	1.2	2.0
Seracare 23	6.8	22.4	1.7	2.1
Seracare 24	13.8	>100	17.2	12.9
Seracare 25	11.8	>100	8.0	7.0

Anti-HBs antibody

Anti-HBc/HBs Mixed Titer Performance Panel

Member ID#	Theranos mIU/mL	Abbott Anti-HBs Architect mIU/mL
PHG203-01	1.00	13.0
PHG203-04	1138	>1000.0
PHG203-05	592.36	>1000.0
PHG203-06	1.25	2.8
PHG203-07	1025.47	687.5
PHG203-08	4.65	20.0
PHG203-09	2.48	51.0
PHG203-10	121.53	175.0
PHG203-11	1.66	0.4
PHG203-12	0.20	0.2
PHG203-13	2.16	0.6
PHG203-14	0.42	0.7

D-Amphetamine

Sample #	AMP reported Conc. (ng/ml)	Target value	Qualitative determination cut-off 500ng/ml	Qualitative determination cut-off 1000ng/ml
Bio Rad C2LO (468)	375	25% lower than cut-off 500ng/ml	NEG	NEG
Bio Rad C3LO (469)	625	25% higher than cut-off 500ng/ml	POS	NEG
Bio Rad S1LO (466)	750	25% lower than cut-off 1000ng/ml	POS	NEG
Bio Rad S2LO (467)	1250	25% higher than cut-off 1000ng/ml	POS	POS
Bio Rad S1 (461)	750	25% lower than cut-off 1000ng/ml	POS	NEG
Bio Rad S2 (462)	1250	25% higher than cut-off 1000ng/ml	POS	POS
Bio Rad S3 (463)	2000	2-fold higher than cut-off 100ng/ml	POS	POS
Bio Rad Neg (460)		negative control	53	NEG

Ethyl glucuronide

Sample ID	Alcohol Consumption in Past 4 Days?	Total Drinks in 4 Days	Time Since last consumption (H)	Conc, ug/mL	
				Mean Conc	CV %
2	No			0.1	8.8
3	No			0.1	29.0
4	Yes	1	48	0.1	5.9
7	No			0.1	6.8
8	Yes	6	14	0.2	12.0
9	No			0.1	11.3
10	Yes	4	17	20.7	24.2
11	Yes	30	12	1.3	17.8
12	Yes	4	15	2.3	26.4
13	Yes	4	20	0.2	6.3
14	Yes	25	20	4.7	51.9
15	No			0.1	7.6
19	No			OORL	
20	No			OORL	
21	No			OORL	
23	No			OORL	
25	Yes	2	42	OORL	
26	Yes	2	45	OORL	
30	Yes	1 54	40	0.1	10.6

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

Helicobacter pylori IgG

Sample	Mean RLU	CV (%)	INOVA	10 ⁴ STDEV			
				Units	BQ	IBL	Theranos
				Units	Units	U/mL	Ab Index
CI01	16629	8	32.5	0.97	1.59	2.3	
CI02	106785	17	128.4	3.02	2.39	14.7	
CI03	121471	8	43.9	0.91	1.98	16.7	
CI04	106361	15	44.0	0.91	1.86	14.6	
CI05	58291	28	61.9	1.94	2.15	8.0	
CI06	109997	9	30.5	1.08	1.95	15.1	
CI07	239170	5	143.3	3.09	2.58	32.9	
CI08	8952	5	43.2	1.26	1.33	1.2	
CI09	273998	8	80.8	2.18	2.49	37.7	
CI10	65144	9	16.8	0.73	1.29	9.0	
CI11	31908	8	33.4	1.19	1.64	4.4	
CI12	347859	14	144.4	3.13	2.62	47.8	
CI13	90265	23	40.7	0.84	1.93	12.4	
CI14	398081	18	145.2	3.11	2.64	54.7	
CI15	163405	5	130.8	3.07	2.50	22.5	
CI16	177040	20	38.1	0.73	2.19	24.3	
CI17	149028	2	45.6	0.88	2.00	20.5	

Sample	Mean RLU	CV (%)	INOVA	10 ⁴ STDEV			
				Units	BQ	IBL	Theranos
				Units	Units	U/mL	Ab Index
CI18	451786	14	145.6	3.23	2.64	62.1	
CI19	35238	42	122.8	2.88	2.11	4.8	
CI20	62668	22	121.7	2.91	2.25	8.6	
CI21	28213	12	83.2	1.96	1.57	3.9	
CI22	97115	23	112.1	2.61	2.13	13.4	
CI23	129145	3	69.9	2.33	2.40	17.8	
CI24	112093	3	42.6	0.82	1.93	15.4	
CI25	364396	7	140.9	3.12	2.66	50.1	
CI26	82713	15	42.7	0.76	1.85	11.4	
CI27	34283	26	64.5	2.05	1.89	4.7	
CI28	148340	22	99.3	2.41	2.23	20.4	
CI29	84895	12	88.4	2.43	2.29	11.7	
CI30	7816	15	27.7	0.77	1.00	1.1	
CI31	93622	14	88.5	1.87	2.07	12.9	
CI32	107219	17	55.3	2.00	2.41	14.7	
CI33	136232	11	41.6	0.69	2.00	18.7	
CI34	245687	15	65.9	2.13	2.50	33.8	

Results reported as: Negative (green), Equivocal (Yellow), or Positive (Red) on the basis of Antibody index for each assay.

Hepatitis B Surface Antigen (HbSAg)


Sample	Theranos Result	ORTHO HBsAg ELISA Testsystem 3	DADE BEHRING Enzygnost HBsAg 5.0	Abbott Murex HBsAg Ver. 3	SORIN ETI-MAK 4	Result	Theranos [S/Co]	CLIA Lab [S/Co]
11000-1	Non-reactive	Non-reactive	Non-reactive	Non-reactive	Non-reactive	Non-reactive	0.874	0.580
11000-2	Non-reactive	Non-reactive	Non-reactive	Non-reactive	Non-reactive	Non-reactive	0.872	0.747
11000-3	Non-reactive	Non-reactive	Non-reactive	Non-reactive	Non-reactive	Non-reactive	0.719	0.747
11000-4	Non-reactive	Non-reactive	Non-reactive	Reactive	Non-reactive	Non-reactive	0.847	0.680
11000-5	Non-reactive	Non-reactive	Non-reactive	Reactive	Non-reactive	Non-reactive	0.747	0.663
11000-6	Non-reactive	Non-reactive	Non-reactive	Reactive	Non-reactive	Non-reactive	0.739	0.753
11000-7	Non-reactive	Reactive	Reactive	Reactive	Reactive	Non-reactive	0.799	0.911
11000-8	Reactive	Reactive	Reactive	Reactive	Reactive	Reactive	1.216	1.070
11000-9	Reactive	Reactive	Reactive	Reactive	Reactive	Reactive	2.967	3.300

Heterophile antibody

Sample ID	Theranos Ab	Osom Mono	Status Mono
1	NEG	NEG	NEG
2	NEG	NEG	NEG
3	NEG	NEG	NEG
4	NEG	NEG	NEG
5	NEG	NEG	NEG
6	NEG	NEG	NEG
7	NEG	NEG	NEG
8	NEG	NEG	NEG
9	NEG	NEG	NEG
10	NEG	NEG	NEG
11	NEG	NEG	NEG
12	NEG	NEG	NEG
13	NEG	NEG	NEG
14	NEG	NEG	NEG
15	NEG	NEG	NEG
16	NEG	NEG	NEG
17	NEG	NEG	NEG
18	NEG	NEG	NEG
19	NEG	NEG	NEG
20	NEG	NEG	NEG
21	NEG	NEG	NEG
22	NEG	NEG	NEG
23	NEG	NEG	NEG
24	NEG	NEG	NEG
25	NEG	NEG	NEG
26	NEG	NEG	NEG
27	NEG	NEG	NEG
28	NEG	NEG	NEG
29	NEG	NEG	NEG


Sample ID	Theranos Ab	Osom Mono	Status Mono
30	NEG	NEG	NEG
31	NEG	NEG	NEG
32	NEG	NEG	NEG
33	NEG	NEG	NEG
34	NEG	NEG	NEG
35	NEG	NEG	NEG
36	NEG	NEG	NEG
37	NEG	NEG	NEG
38	NEG	NEG	NEG
39	BOR	NEG	NEG
40	NEG	POS	NEG
41	POS	POS	POS
42	POS	POS	POS
43	POS	POS	POS
44	POS	POS	POS
45	POS	POS	POS
46	POS	POS	POS
47	POS	POS	POS
48	POS	POS	POS
49	POS	POS	POS
50	POS	POS	POS
51	POS	POS	POS
52	POS	POS	POS
53	POS	POS	POS
54	POS	POS	POS
55	POS	POS	POS
56	POS	POS	POS
57	POS	POS	POS
58	POS	POS	POS

Human chorionic gonadotropin (hCG)



Human IgD

Clinical samples Siemens vs Theranos result



Anti-HSV1 IgG

Liaison HSV2 Result	Sample ID	Liaison Result	EuroImmun Result	Theranos HSV1 Result
Negative	BR101	Positive	Positive	Positive
Negative	BR103	Positive	Positive	Positive
Negative	BR104	Positive	Positive	Positive
Negative	BR105	Positive	Positive	Positive
Negative	BR106	Positive	Positive	Positive
Negative	BR107	Positive	Positive	Positive
Negative	BR109	Positive	Positive	Positive
Negative	BR111	Positive	Positive	Positive
Negative	BR112	Positive	Positive	Positive
Negative	BR114	Positive	Positive	Positive
Negative	BR115	Positive	Positive	Positive
Negative	BR116	Positive	Positive	Positive
Negative	BR117	Positive	Positive	Positive
Negative	SC01	Positive	Positive	Positive
Negative	SC02	Positive	Positive	Positive
Negative	SC07	Positive	Positive	Positive
Negative	SC10	Positive	Positive	Positive
Negative	SC16	Positive	Positive	Positive
Negative	SC17	Positive	Positive	Positive
Negative	SC18	Positive	Positive	Positive
Negative	Z01	Positive	Positive	Positive

Liaison HSV2 Result	Sample ID	Liaison Result	EuroImmun Result	Theranos HSV1 Result
Negative	Z01	Positive	Positive	Positive
Negative	Z06	Positive	Positive	Positive
Negative	Z11	Positive	Positive	Positive
Negative	Z13	Positive	Positive	Positive
Negative	Z14	Positive	Positive	Positive
Negative	Z20	Positive	Positive	Positive
Negative	Z26	Positive	Positive	Positive
Negative	Z27	Positive	Positive	Positive
Negative	Z28	Positive	Positive	Positive
Negative	Z29	Positive	Positive	Positive
Negative	Z30	Positive	Positive	Positive
Negative	Z31	Positive	Positive	Positive
Negative	Z32	Positive	Positive	Positive
Negative	Z34	Positive	Positive	Positive
Negative	Z35	Positive	Positive	Positive
Negative	Z36	Positive	Positive	Positive
Negative	MT01	Positive	Positive	Positive
Negative	MT21	Positive	Positive	Positive
Positive	BR108	Positive	Positive	Positive
Positive	BR118	Positive	Positive	Negative
Positive	BR201	Positive	Positive	Positive
Positive	BR206	Positive	Positive	Positive

Anti-HSV2

Liaison HSV1 Result	Sample ID	Liaison HSV2 Result	Eurolimmun HSV2 Result	Theranos HSV2 Result
Negative	BR113	Positive	Positive	23.34
Negative	BR202	Positive	Positive	29.79
Negative	BR203	Positive	Positive	7.94
Negative	BR204	Positive	Positive	2.63
Negative	BR207	Positive	Positive	15.77
Negative	BR208	Positive	Positive	24.24
Negative	BR209	Positive	Positive	14.48
Negative	BR211	Positive	Positive	24.93
Negative	BR213	Equivocal	Negative	0.74
Negative	BR215	Positive	Positive	13.33
Negative	Z16	Positive	Equivocal	5.14
Negative	Z21	Positive	Positive	9.96
Negative	Z22	Positive	Equivocal	5.22
Negative	Z23	Positive	Equivocal	3.63
Negative	F03	Positive	Positive	13.34
Negative	SC03	Positive	Positive	10.56
Negative	SC04	Positive	Positive	36.49
Negative	SC05	Positive	Positive	14.44
Negative	SC11	Positive	Positive	13.15
Negative	SC15	Positive	Positive	4.80
Negative	MT02	Positive	Positive	18.50
Negative	MT06	Positive	Equivocal	2.82
Negative	MT07	Positive	Equivocal	0.78
Negative	MT08	Positive	Positive	23.58
Negative	MT13	Positive	Positive	17.13
Negative	MT16	Positive	Positive	3.54
Negative	MT17	Positive	Positive	22.53
Negative	MT24	Positive	Equivocal	3.33
Positive	BR108	Positive	Positive	17.76
Positive	BR110	Positive	Positive	10.53
Positive	BR118	Positive	Positive	21.22
Positive	BR201	Positive	Positive	14.83
Positive	BR206	Positive	Positive	18.70
Positive	BR210	Positive	Positive	8.29
Positive	BR212	Positive	Positive	17.80
Positive	BR214	Positive	Positive	8.72

Influenza A – Qualitative EIA


TYPE	SAMPLE ID	Theranos		REMEL (FDA)
		Ab index	Result	Result
NORMAL CLINICALS	1	0.02	NEG	NEG
	6	0.02	NEG	NEG
	7	0.02	NEG	NEG
	8	0.02	NEG	NEG
	10	0.02	NEG	NEG
	11	0.01	NEG	NEG
	12	0.02	NEG	NEG
	13	0.01	NEG	NEG
	15	0.02	NEG	NEG
	16	0.04	NEG	NEG
	17	0.02	NEG	NEG
	18	0.02	NEG	NEG
	2	0.32	NEG	NEG
	3	0.23	NEG	NEG
	4	0.20	NEG	NEG
	9	0.05	NEG	NEG
	14	0.29	NEG	NEG
	19	0.95	NEG	NEG
REMEL	Positive control swab	2.66	POS	POS
Zeptometric QC panel	Flu A POS	1.27	POS	POS
Flu A	Brisbane/10/07	2.58	POS	POS
Flu A	Solomon Islands/03/2006	3.25	POS	POS
Flu A	New Caledonia/20/99	2.57	POS	POS
Flu A	Brisbane/59/07	5.16	POS	POS
NIBSC STANDARDS FLU B	Panama/45/90	0.06	NEG	
	Influenza Antigen B-Johannesburg	0.06	NEG	
	Influenza Antigen B-Guangdong	0.08	NEG	
	Influenza Antigen B/Yamanashi/166/98	0.11	NEG	
	Influenza Antigen B/Malaysia/2506/2004	0.02	NEG	
	Influenza Antigen B/Harbin/7/94	0.06	NEG	
	B/Florida 4/2006	0.04	NEG	
NIBSC STANDARDS FLU A	Influenza Antigen A/California/7/2009-H1N1	6.88	POS	
	Influenza Antigen A/HongKong/1073/99 (H9N2)	8.56	POS	
	Influenza Antigen A/Cambodia/R0405050/2007 (H5N1)	6.02	POS	
	Influenza Antigen A/mallard/England/727/2006 (H2N3)	5.70	POS	
	Influenza Antigen A/New York/107/2003 (H7N2) (NIBRG-109)	7.26	POS	
	Influenza Antigen A/New York/55/2004 (H3N2) (NYMC X-157)	6.54	POS	

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

Influenza B – Qualitative EIA

TYPE	SAMPLE ID	Theranos		REMEL (FDA)
		Ab index	Result	Result
NORMAL CLINICALS	1	0.02	NEG	NEG
	6	0.03	NEG	NEG
	7	0.02	NEG	NEG
	8	0.02	NEG	NEG
	10	0.04	NEG	NEG
	11	0.02	NEG	NEG
	12	0.03	NEG	NEG
	13	0.01	NEG	NEG
	15	0.02	NEG	NEG
	16	0.06	NEG	NEG
	17	0.02	NEG	NEG
	18	0.03	NEG	NEG
	2	0.11	NEG	NEG
	3	0.03	NEG	NEG
	4	0.04	NEG	NEG
	9	0.05	NEG	NEG
	14	0.36	NEG	NEG
	19	0.79	NEG	NEG
REMEL	FDA Pos B Swab	10.78	POS	POS
Zeptometric QC panel	Flu A POS	0.02	NEG	NEG
Flu A	Brisbane/10/07	0.03	NEG	NEG
Flu A	Solomon Islands/03/2006	0.01	NEG	NEG
Flu A	New Caledonia/20/99	0.02	NEG	NEG
Flu A	Brisbane/59/07	0.03	NEG	NEG
NIBSC STANDARDS FLU B	Panama 45/90	14.38	POS	POS
	Influenza Antigen B-Johannesburg	2.58	POS	POS
	Influenza Antigen B-Guangdong	21.28	POS	POS
	Influenza Antigen B/Yamanashi/166/98	6.05	POS	POS
	Influenza Antigen B/Malaysia/2506/2004	7.53	POS	POS
	Influenza Antigen B/Harbin/7/94	18.21	POS	POS
	B/Florida 4/2006	19.27	POS	POS
NIBSC STANDARDS FLU A	Influenza Antigen A/California/7/2009-H1N1	0.36	NEG	NEG
	Influenza Antigen A/HongKong/1073/99 (H9N2)	0.50	NEG	NEG
	Influenza Antigen A/Cambodia/R0405050/2007 (H5N1)	0.61	NEG	NEG
	Influenza Antigen A/mallard/England/727/2006 (H2N3)	0.50	NEG	NEG
	Influenza Antigen A/New York/107/2003 (H7N2) (NIH RG-109)	0.39	NEG	NEG
	Influenza Antigen A/New York/55/2004 (H3N2) (NYMC X-157)	0.12	NEG	NEG
Zeptometric Panel Flu B	Lee/40	11.31	POS	POS
Flu B	Florida/02/2006	2.57	POS	POS
Flu B	Brisbane/33/2008	12.36	POS	POS
Flu B	Panama/45/90	5.93	POS	POS
Flu B	Panama/45/90	4.64	POS	POS

Immunoglobuline (IgE)




Anti-Jo-1

				10⁶ STDEV
	Innova	Immco	IBL Int.	Theranos
Sample	Units	Result (EU/mL)	Ratio	Ab Index
CI01	0	6	0.1	0.6
CI02	1	7	0.1	0.7
CI03	32	126	4.4	5.7
CI04	127	177	6.6	18.1
CI05	111	169	6.6	9.6
CI06	5	8	0.1	0.7
CI07	19	4	0.1	0.0
CI08	3	10	0.4	0.6
CI09	8	13	0.3	0.7
CI10	1	14	0.2	0.7
CI11	2	12	0.3	0.0
CI12	1	7	0.1	0.0
CI13	0	5	0.1	0.0
CI14	1	3	0.1	0.0
CI15	1	20	0.2	0.3
CI16	0	8	0.1	0.0
CI17	165	175	6.2	16.6
CI18	1	8	0.2	0.0
CI19	50	138	4.4	8.1
CI20	9	10	0.2	0.0
CI21	107	175	6.3	8.9
CI22	8	8	0.1	0.0
CI23	3	10	0.4	0.0
CI24	5	14	0.3	0.5
CI25	23	2	0.1	0.4
CI35	107	201	7.9	2.5
CI36	69	158	6.5	2.3
CI37	110	203	8.1	2.6
CI38	108	208	8.2	2.7
CI39	109	201	8.2	3.0

Results reported as Negative (green) or Positive (Red) on the basis of Antibody index.

Luteinizing Hormone (LH)




Lyme disease Antibody

Sample	US Biological Ab Index	Immunetics Ab Index	Theranos Ab Index
10915791	3.49	4.43	2.35
10919479	0.54	0.25	0.32
10919480	1.17	0.43	0.53
10919776	3.96	4.71	1.57
10919778	1.66	2.09	1.13
10919847	5.15	9.18	113.95
10919848	0.58	0.26	0.24
10919849	0.61	0.22	0.21
10919906	2.06	0.58	0.33
10919907	2.72	8.36	24.21
10919908	3.79	9.06	14.76
10919970	3.59	0.23	1.01
10919972	3.33	4.71	5.45
10919973	2.14	7.00	5.01
10920987	4.33	9.10	36.45
10921154	1.17	0.24	0.86
10921214	0.89	0.33	0.25
10921215	2.31	0.55	0.50
10921305	2.03	2.81	1.26
10921353	0.83	0.29	0.38
10924272	4.29	4.96	4.53
10924273	0.83	0.20	0.26
10924274	3.21	0.20	0.45
10924275	0.47	0.14	0.14
10924276	1.81	2.65	1.71
10924347	3.54	7.24	6.07
10924348	0.76	0.17	0.24
10924351	3.38	6.55	6.40
10924352	4.07	7.74	17.42
10924353	1.84	2.12	1.21


Results reported as:
Negative (green),
Equivocal (Yellow), or
Positive (Red) on the
basis of Antibody
index for each assay.

Measles IgG

Measles IgG assay: Theranos vs. GSK correlation
: N=30 post vaccinated samples: Excluding
Sample #27



Clinical sample correlation: Theranos vs.
Serion ELISA



Measles IgM

Samples	Theranos Ab Index	Theranos Result	GSD Measles Measles IgM Kit
1	0.48	NEG	NEG
2	0.34	NEG	NEG
3	0.32	NEG	NEG
4	0.37	NEG	NEG
5	0.37	NEG	NEG
6	0.40	NEG	NEG
7	0.31	NEG	NEG
8	0.44	NEG	NEG
9	0.42	NEG	NEG
10	0.44	NEG	NEG
11	0.32	NEG	NEG
12	0.32	NEG	NEG
13	0.33	NEG	NEG
14	0.31	NEG	NEG
15	0.33	NEG	NEG
16	0.33	NEG	NEG
17	0.32	NEG	NEG
18	0.38	NEG	NEG
19	0.43	NEG	NEG
20	0.32	NEG	NEG
21	0.33	NEG	NEG
22	0.24	NEG	NEG
23	0.30	NEG	NEG
24	0.31	NEG	NEG
25	0.29	NEG	NEG
26	0.30	NEG	NEG
27	0.33	NEG	NEG
28	0.31	NEG	NEG
29	0.33	NEG	NEG
30	0.20	NEG	NEG
33	0.29	NEG	NEG
34	0.27	NEG	NEG
35	0.38	NEG	NEG
36	0.33	NEG	NEG
42	22.77	POS	POS

Methadone

Sample ID (Urine)	CLIA Advia Results ng/ml	Theranos Result		% Recovery
		Mean ng/mL		
U1	1286	OORH	Positive	OORH
U2	913	OORH	Positive	OORH
U3	700	898	Positive	128
U4	671	762	Positive	114
U5	551	575	Positive	104
U6	386	416	Positive	108
U7	289	248	Negative	86
U8	271	284	Negative	105
U9	217	216	Negative	100
U10	199	236	Negative	119
U11	685	727	Positive	106
U12	587	577	Positive	98
U13	446	471	Positive	105
U14	454	370	Positive	81
U15	406	372	Positive	92
U16	334	321	Positive	96
U17	184	196	Negative	107
U18	157	170	Negative	108
U19	151	163	Negative	108
U20	130	91	Negative	70
U21		749	Positive	75
U22		355	Positive	118

Sample ID	Nominal [Methadone] ng/ml	Inter-Cartridge		Inter- Cartridge		% Recovery	Theranos Result
		RLU Mean	CV%	Conc. ng/ml	CV%		
P1	1200	6684	7.9	1048	13.8	87	Positive
P2	950	8297	11.0	742	15.2	78	Positive
P3	700	10397	8.4	534	10.0	76	Positive
P4	550	12294	6.8	433	8.1	79	Positive
P5	375	17688	5.4	295	5.3	79	Positive
P6	225	27217	10.2	198	10.2	88	Positive
P7	155	41133	31.4	140	26.4	90	Positive
P8	65	88034	13.4	61	15.7	93	Positive
P9	40	100276	2.5	50	3.6	124	Positive
P10	15	172327	16.8			OORL	Negative


Opiates

Sample [Urine]	Theranos		CLIA		
	[ng/mL]	[ng/mL]	Interpretation	[ng/mL]	Interpretation
Heroin	200	158	NEG	150	NEG
	650	693	POS	535	POS
	32	35	NEG	40	NEG
	17	12	NEG	NA	NA
	5	5	NEG	NA	NA
6-Acetylmorphine	20	26	NEG	24	NEG
	45	46	NEG	55	NEG
	677	625	POS	563	POS
	10	8	NEG	NA	NA
	301	335	POS	256	NEG
Codeine	75	84	NEG	94	NEG
	10	8	NEG	NA	NA
	15	13	NEG	33	NEG
	505	538	POS	1547	POS
	29	34	NEG	25	NEG
Biorad C2 (low opiate)	225	261	NEG	727	POS
Biorad C3 (low opiate)	375	487	POS	1426	POS
Biorad C1	120	143	NEG	271	NEG
Biorad C2	1500	1129	POS	NA	NA
Biorad C3	2500	1538	POS	NA	NA
Biorad C4	4000	1821	POS	NA	NA
Biorad N (negative)	75	1	NEG	NA	NA
Sunnylab opiate	3040	3040	POS	NA	NA
Biorad QN	50	3	NEG	NA	NA
Biorad QP	2904	694	POS	3404	POS
Biorad S1S	1500	454	POS	1259	POS
Biorad S2S	2500	615	POS	2273	POS
Biorad S1E (low opiate)	225	81	NEG	216	NEG


Parvovirus IgG

	Inter Cartridges		Theranos	Commercial Kit		Inter Cartridges		Theranos	Commercial Kit
Sample #	Ave RLU	%CV	Ab Index 10X	Ab Index	Sample #	Ave RLU	%CV	Ab Index 10X	Ab Index
1	31383	12	13.672	Positive	16	42529	21	18.528	Positive
2	43388	14	18.902	Positive	17	75966	16	33.095	Positive
3	26851	19	11.698	Positive	18	97569	13	42.506	Positive
4	7779	91	5.415	Positive	19	16431	16	7.158	Positive
5	56796	24	24.743	Positive	20	14844	25	6.467	Positive
6	51658	19	22.505	Positive	21	8411	18	3.664	Positive
7	88529	15	38.568	Positive	22	9534	16	4.154	Positive
8	31634	21	13.781	Positive	23	2920	24	1.272	Positive
9	22517	17	9.810	Positive	24	5005	20	2.180	Positive
10	15221	26	6.631	Positive	25	3755	13	1.636	Positive
11	53161	21	23.160	Positive	26	11917	21	5.192	Positive
12	19963	25	8.697	Positive	27	5569	12	2.426	Positive
13	75049	13	32.695	Positive	28	13348	18	5.815	Positive
14	52442	20	22.847	Positive	29	15502	18	6.753	Positive
15	108177	6	47.128	Positive ⁷²	30	17819	17	7.763	Positive


Progesterone



Prolactin



Total PSA



RF IgG antibody


Set Type	Sample#	Theranos S/Co	DRG Mean Conc (U/ml)	Alpco Mean Conc (U/ml)
ProMedDx	1	13.5	993	2143
	2	12.1	464	169
	3	20.2	298	121
	6	18.7	624	81
	9	10.8	366	70
	10	48.3	27	73
	11	22.1	357	987
	12	8.2	234	160
	13	12.3	35	94
	14	8.9	407	72
	15	3.9	505	217
	16	6.0	550	115
	17	2.3	61	221
	18	11.7	902	90
	23	1.7	113	129
	26	1.2	311	102
	31	16.7	540	120
	33	7.9	121	145
	37	4.6	128	97
	39	1.7	125	312
Bioreclamation (Normal Seras)	M11	0.1	19	69
	M12	0.1	12	29
	M13	0.1	15	57
	M14	0.1	16	31
	M15	0.1	8	30
	F11	0.1	15	52
	F12	0.1	10	24
	F13	0.1	16	51
	F14	0.1	17	54
	F15	0.1	10	24
	Set 2	1	0.1	6
		2	0.1	5
		3	0.1	5
		5	0.1	6
		6	0.1	8
		20	0.1	5
		21	0.1	3
		22	0.8	5
		29	0.1	5
				35

RF IgM antibody

Set Type	Sample ID	ALPCO	INOVA	DRG	Theranos
		Mean Conc(IU/mL)	Mean Conc(IU/mL)	Mean Conc(IU/mL)	Mean Conc (IU/mL)
Positive	19	166	573	347	190
	27	741	2191	815	1397
	38	159	291	25	1885
	40	22	20	25	OORL
	41	602	766	1187	1063
	42	858	1210	839	1577
	43	232	381	778	354
	44	78	187	19	891
	45	597	993	5603	1941
	46	484	988	1217	987
	47	273	914	2110	708
	48	316	914	722	744
	49	1048	2982	942	1652
	50	580	1071	97	1266
	51	258	632	221	545
	52	277	314	1728	645
	53	912	1423	980	1304
	54	179	128	OORL	835
	55	171	277	135	365
	56	93	116	100	125
	57	101	1805	956	NES
	58	552	948	348	1448
	59	413	783	863	900
	60	508	1561	328	1159

Set Type	Sample ID	ALPCO	INOVA	DRG	Theranos
		Mean Conc(IU/mL)	Mean Conc(IU/mL)	Mean Conc(IU/mL)	Mean Conc (IU/mL)
Negative	B1	0.9	0.6	4.0	OORL
	B2	0.9	0.2	OORL	OORL
	B3	1.5	1.9	OORL	OORL
	B4	2.0	1.1	7.0	OORL
	B5	1.2	1.0	3.2	OORL
	B6	1.0	0.6	OORL	OORL
	B7	0.7	0.2	OORL	OORL
	B8	0.7	0.3	OORL	OORL
	B9	8.0	9.3	13.3	OORL
	B10	1.6	0.9	7.2	OORL
	B11	1.4	1.1	OORL	OORL
	B12	3.4	2.9	0.6	OORL

Rubella Antibody



Rubella IgM

Samples	Theranos Ab Index	Theranos Result	Siemens Immulite 2000
1	0.12	NEG	NEG
2	0.13	NEG	NEG
3	0.10	NEG	NEG
4	0.29	NEG	NEG
5	0.54	NEG	NEG
6	0.26	NEG	NEG
7	0.19	NEG	NEG
8	0.22	NEG	NEG
9	0.12	NEG	NEG
10	0.13	NEG	NEG
11	0.07	NEG	NEG
12	0.06	NEG	NEG
13	0.06	NEG	NEG
14	0.15	NEG	NEG
15	0.07	NEG	NEG
16	0.29	NEG	NEG
17	0.09	NEG	NEG
18	0.18	NEG	NEG
19	0.10	NEG	NEG
20	0.09	NEG	NEG

Samples	Theranos Ab Index	Theranos Result	Siemens Immulite 2000
21	0.87	NEG	NEG
22	0.08	NEG	NEG
23	0.23	NEG	NEG
24	0.09	NEG	NEG
25	0.10	NEG	NEG
26	0.11	NEG	NEG
27	0.13	NEG	NEG
28	0.07	NEG	NEG
29	0.25	NEG	NEG
30	0.09	NEG	NEG
31	0.32	NEG	POS
32	2.95	POS	POS
33	3.36	POS	POS
34	2.11	POS	POS
35	3.70	POS	POS
36	3.91	POS	POS
37	6.06	POS	POS
38	1.59	POS	POS
39	3.56	POS	POS
40	7.48	POS	POS

Anti-Scl 70 antibody

Sample ID	Human Test Samples			ANTIBODY INDEX			
	Matrix	Species	Strain	Theranos	INOVA	IMMCO	IBL
CLN1	Serum	Normal	N/A	0.04	4	5	0.15
CLN2	Serum	Normal	N/A	0.04	4	2	0.07
CLN3	Serum	Normal	N/A	0.06	4	6	0.16
CLN4	Serum	Normal	N/A	0.09	4	2	0.13
CLN5	Serum	Normal	N/A	0.07	5	3	0.10
CLN6	Serum	Normal	N/A	0.06	3	4	0.16
CLN7	Serum	Normal	N/A	0.05	4	7	0.3
CLN8	Serum	Normal	N/A	0.11	5	5	0.15
CLN9	Serum	Normal	N/A	0.06	3	14	0.08
CLN10	Serum	Normal	N/A	0.05	3	6	0.12
CLN11	Serum	Normal	N/A	0.59	9	33	0.29
CLN12	Serum	Normal	N/A	0.04	5	5	0.22
CLN13	Serum	Normal	N/A	0.05	4	2	0.18
CLN14	Serum	Normal	N/A	0.04	11	174	3.06
CLS1	Serum	Autoimmune	Scleroderma	7.35	9	186	0.33
CLS2	Serum	Autoimmune	Scleroderma	0.72	19	70	0.73
CLS9	Serum	Autoimmune	Scleroderma	1.60	100	213	3.76
SS10	Serum	Autoimmune	Sjogren	1.02	5	91	0.19
SCL06	Serum	Autoimmune	Scleroderma	0.95	23	83	1.78
SCL16	Serum	Autoimmune	Scleroderma	6.35	3	194	3.07
SCL25	Serum	Autoimmune	Scleroderma	3.59	141	267	5.25
SCL29	Serum	Autoimmune	Scleroderma	3.04	124	257	2.66
SCL38	Serum	Autoimmune	Scleroderma	0.24	1	69	0.10
SCL39	Serum	Autoimmune	Scleroderma	3.33	146	293	5.58
SCL40	Serum	Autoimmune	Scleroderma	4.50	139	280	5.78

Samples are considered to be positive, borderline, or negative for Scl-70 antibodies if their Ab Indices are found to be greater than 1.1, between 0.9 and 1.1, or less than 0.9, respectively.

Anti-SSA antibody


	Innova	Immco	IBL Int.	Theranos
Sample	Units	Result (EU/mL)	Ratio	Ab Index
ScL01	42.1	53	0.28	5.0
SCL02	3.9	4	0.38	0.6
ScL03	8.5	9	0.16	0.6
ScL04	17.8	10	0.42	0.4
ScL05	3.8	3	0.44	0.3
ScL06	3.7	2	0.24	0.5
ScL07	4.5	2	0.23	0.3
ScL08	3.7	2	0.22	0.4
ScL09	5.1	5	0.33	0.3
ScL10	17.4	8	0.29	1.1
SL01	102.9	78	5.51	212.8
SL02	4.7	5	0.30	0.5
SL03	3.8	3	0.23	0.4
SL04	70.8	66	5.50	2.5
SL05	3.6	2	0.58	0.3
SL06	59.5	49	5.29	0.9
SL07	4.5	7	0.24	0.6
SL08	86.1	69	5.32	1.7
SL09	3.6	11	0.67	0.6
SL10	3.7	3	0.19	0.4
Sjo01	108.6	57	5.87	33.2
Sjo02	3.5	3	0.23	0.3
Sjo03	125.6	74	6.22	373.3
Sjo04	24.2	57	2.45	14.7
Sjo05	96.9	78	5.61	203.2

The following criteria was applied to categorize the result as positive (red), negative (green) or borderline (yellow).


Ab Index > 1.1
Ab Index > 0.9, < 1.1
Ab Index < 0.9

Sex Hormone Binding Globulin (SHBG)

Clinical Correlation: Theranos vs Siemens Immulite 2000




Total Testosterone




Anti-Thyroglobulin Antibody (IgG)

Clinical Correlation to Reported Siemens Immulite Results, 18 In-Range Samples




Thyroglobulin (antigen)


Theranos vs Orgentec




Theranos vs Siemens Immulite 2000




Thyroid stimulating hormone (TSH)




Thyroid peroxidase Antibody (TPO)




Total Thyroxine (TT4)



Total Triiodothyronine (TT3)



T3 Uptake




The T Uptake % is used to compute the Free Thyroxine Index (FTI) from the Total T4

Transferrin


Sample #			Theranos TRF System 3.0				
			Signal (RLU)		[TRF, back-calculated], mg/dL		% Diff from Target
			SIEMENS ADVIA		Inter-Cartridge		SIEMENS ADVIA
mg/dL	Mean	%CV	Mean	%CV	mg/dL		
1 AD08	293	85703	21	279	18	-4.9	
2 AD10	239	58676	14	205	11	-14.0	
3 AD18	312	88990	19	287	16	-7.9	
4 P7	201	37947	20	148	15	-26.5	
5 P11	414	114683	11	358	10	-13.6	
6 P12	333	93325	25	299	21	-10.1	
7 L3	215	46044	20	171	15	-20.7	
8 L4	290	64780	36	222	28	-23.6	
9 6-BRH539339	259	60136	7	210	6	-19.1	
10 10-BRH539334	314	86289	8	280	6	-10.8	
11 P1	283	71893	9	241	7	-14.8	
12 P3	232	56759	32	212	21	-8.6	
13 BRH460987	345	100216	21	318	18	-7.8	
14 BRH460980	269	56909	24	200	18	-25.5	
15 PD6	248	52493	27	188	21	-24.1	
16 PD7	115	21820	28	99	20	-14.2	
17 PD8	276	82295	21	269	18	-2.4	
18 PD9	229	49823	24	163	30	-28.8	
19 PD10	147	26214	25	113	17	-23.3	
N/A IRMM	236	54947	15	195	12	-17.2	
20 AD03	312	64128	27	234	17	-25.1	
21 AD05	247	49099	33	187	23	-24.3	
22 AD07	320	77237	26	256	21	-20.1	
23 AD11	208	40937	35	156	26	-25.1	
24 7-BRH539338	299	62332	41	215	32	-28.1	
25 8-BRH539336	300	39212	20	151	15	-49.5	
26 9-BRH539340	239	72963	16	244	13	2.1	
27 PD11	284	42540	39	160	29	-43.6	

Varicella Zoster virus - IgG



Vitamin B12

B12 Assay: Clinical Sample Correlation



VZV – IgG Precision

Inter-Analyzer Precision, Concentration (IU/mL)

Analyzer	Tip 1	Tip 2	Intra-Cartridge			Inter-Cartridge	
			Mean	CV %	% Recovery	Mean	CV %
1	0.26	0.27	0.26	4.1	86	0.31	7.2
2	0.34	0.35	0.35	1.1	113		
3	0.30	0.36	0.33	11.4	108		
4	0.26	0.28	0.27	4.8	89		
5	0.30	0.31	0.30	2.1	99		
6	0.28	0.27	0.28	3.6	90		
7	0.29	0.31	0.30	2.7	98		
8	0.32	0.32	0.32	0.3	104		
9	0.30	0.28	0.29	6.1	95		
10	0.32	0.33	0.32	1.0	106		
11	0.28	0.27	0.27	0.8	89		
12	0.30	0.32	0.31	4.4	100		
13	0.29	0.32	0.31	7.5	100		
14	0.35	0.32	0.34	4.8	109		
15	0.35	0.33	0.34	4.0	111		
16	0.31	0.31	0.31	1.2	101		
17	0.29	0.33	0.31	8.8	100		
18	0.31	0.25	0.28	16.8	91		
19	0.32	0.29	0.30	6.3	99		
20	0.29	0.31	0.30	5.9	98		
21	0.30	0.32	0.31	4.3	100		
22	0.33	0.31	0.32	4.7	104		
23	0.33	0.32	0.32	1.9	105		
24	0.33	0.31	0.32	5.0	103		

VZV – IgG Precision

Precision and Accuracy for 3 Days/Lots, Concentration (IU/mL)

N = 3 cartridges per point


[VZV IgG] IU/mL	Day/Lot 1			Day/Lot 2			Day/Lot 3			Inter-Lot		
	Mean Conc.	CV %	% Rec.									
11.00	9.97	22.3	91	11.40	9.5	104	9.83	33.6	89	10.47	18.9	95
5.50	4.65	2.4	85	6.65	13.6	121	6.09	15.9	111	5.94	18.5	108
2.75	2.33	15.3	85	3.07	7.3	112	2.65	8.1	96	2.68	14.9	98
1.38	1.27	13.5	92	1.46	10.6	106	1.46	12.3	107	1.40	12.6	101
0.69	0.65	11.1	94	0.75	6.7	109	0.71	6.9	104	0.70	9.6	102
0.34	0.32	7.8	93	0.34	13.7	98	0.35	4.0	103	0.34	9.2	98
0.17	0.17	9.2	99	0.18	9.7	105	0.17	1.9	98	0.17	7.5	101
0	OORL	-	-									
Positive Control	4.25	8.6	-	3.70	11.7	-	4.25	9.3	-	4.05	11.0	-
Negative Control	OORL	-	-									

West Nile IgM

Samples	Recombinant WNV Ag		NCA		Theranos		Focus		Panbio	
	Mean	CV	Mean	CV	S/co	Result	S/o	Result	S/o	Result
PWN 901-1.5 Member 1	2391	13	297	25	0.23	Negative	0.00	Negative	0.20	Negative
PWN 901-1.5 Member 2	2481	23	319	4	0.24	Negative	0.00	Negative	0.20	Negative
PWN 901-1.5 Member 3	2380	21	417	14	0.22	Negative	0.00	Negative	0.20	Negative
PWN 901-1.5 Member 4	3144	9	398	22	0.30	Negative	0.10	Negative	0.20	Negative
PWN 901-1.5 Member 5	165647	20	303	9	18.34	Positive	4.10	Positive	2.00	Positive


Human IgM

clinical and control correlation
(40 samples + 8 controls)




PCP : Calibration Verification

Correlation to GC-MS Result, Urine Controls




general chemistry


Routine Test Validations Demonstrate High Correlation Coefficients Across Clinical Range




Acetaminophen in Plasma (Lithium Heparin)



Activate Partial Thromboplastin Time (aPTT) Plasma spiked with Heparin




Theranos
results




Commercially available
POC device results

PT results – Clinical samples of patients on Coumadin




Alanine Transaminase (ALT)

Clinical Correlation
(N= 13 Samples)




Albumin




Aldolase (Serum/Plasma)


Clinical Correlation for Theranos Aldolase Assay vs Caldon Biotech Aldolase Assay: All 15 Samples




Alkaline Phosphatase (ALP)




Alpha Amylase (Plasma)



Ammonia (Plasma)




Aspartate Aminotransferase




Direct Bilirubin (Plasma)

Clinical Sample Correlation




Total Bilirubin (Plasma)

Clinical Sample Correlation



Bilirubin (Urine)




Blood Typing: - ABO & RhD

Blood Type	Test Results		Blood Type	Test Results	
	Reference	Theranos		Reference	Theranos
EDTA	O +	O +	EDTA	AB -	AB -
EDTA	O +	O +	EDTA	A -	A -
EDTA	B +	B +	EDTA	O -	O -
EDTA	O +	O +	Heparin/FS	AB +	AB +
EDTA	A +	A +	Heparin/FS	A -	A -
EDTA	A +	A +	Heparin/FS	B +	B +
EDTA	O -	O -	Heparin/FS	B +	B +
EDTA	AB +	AB +	Heparin/FS	O -	O -
EDTA	O +	O +	Heparin/FS	A +	A +
EDTA	A +	A +	Heparin/FS	AB -	AB -
EDTA	A +	A +	Heparin/FS	O +	O +
EDTA	O +	O +	Heparin/FS	B +	B +

Ref: Theranos Folder: J:\Experiment Log\E0700 - E0799\E0705\ABO-Rh Typing\ABO minitips images




Blood Urea Nitrogen (BUN)




Calcium

Clinical Sample Correlation




Carbon Dioxide


N = 16 samples




Total Cholesterol



Urine Based Assays – Creatinine, Chloride




Creatinine (Urine)



Chloride

N = 41 clinical and spiked samples




Creatine Kinase


Clinical Sample Correlation for Theranos and Teco CK Assays: 12 Samples

$$y = 0.93x + 38.22$$


$R^2 = 0.96$




Creatinine (Plasma)



eGFR




Formate (Plasma)



General Chemistries


GGT

Clinical Correlation between Theranos GGT Assay and BioQuant Soluble Reagent GGT Kit: 17 Samples







Ethanol


Clinical Correlation between Theranos Ethanol Assay and Promeddx: 25 Samples



General Chemistries




Glucose




Hemoglobin

N = 36 clinical samples


Clinical Correlation: Teco versus Theranos
(Whole Blood Samples)




Heparin anti-factor Xa



Intact Parathyroid Hormone (iPTH)




Iron (Serum)




Total Iron Binding Capacity (TIBC)

N = 16 samples

Clinical correlation



Hexosaminidase Total & A in Plasma




Lactate Dehydrogenase (LDH)


Theranos vs BioAssay LDH Calculated Activity for
Clinicals: 17 Samples

$$y = 0.99x + 4.05$$

$R^2 = 0.93$




Direct LDL-Cholesterol




Lipase


Clinical Correlation




Phosphate



Potassium




Pyruvate (Plasma)




Sodium (Urine)

Sodium Urine Clinical Correlation (N=60 Samples)




Sulfate (Urine)




Streptozyme Assay Summary

		Theranos	
		Negative	Positive
Reference Negative	17	1*	94% Specificity
	0	20	100% Sensitivity


Triglycerides (Plasma)




Uric Acid




VLDL-Cholesterol



Zinc (Plasma/Serum)






Zinc Protoporphyrin






cytometry

Total WBC, RBC and Platelet Counts: correlation between Theranos platform and hematology analyzer



Correlation of WBC-differential assay between Theranos platform and reference hematology analyzer

All numbers are WBC cell type proportions expressed as percentages of total WBC



Epithelial Cells – Urine Microscopy

Epithelial Cells

		none	rare	few	mod	many	
		none	2	0	0	0	0
REF Theranos	none	10	8	0	0	0	
	rare	1	1	2	0	0	
	few	0	2	0	0	0	
	mod	1	0	0	0	0	
	many	0	0	0	0	0	

WBC – Urine Microscopy

WBCs

REF		none	rare	occ	few	mod	many
Theranos							
none	2	2	0	0	0	0	
rare	1	13	1	0	0	0	
occ	0	3	0	0	0	0	
few	0	2	0	0	0	0	
mod	0	1	0	0	0	0	
many	0	0	0	0	0	2	

RBC – Urine Microscopy


RBCs

		none	rare	occ	few	mod	many
		none	rare	occ	few	mod	many
Theranos	none	1	0	0	1	0	0
	rare	4	6	1	1	0	0
	occ	0	4	1	0	0	0
	few	0	3	2	0	0	0
	mod	0	0	0	0	0	0
	many	0	1	1	0	1	0


nucleic acid amplification

NAA detection (E. Coli O157)


Negative



Positive



Dengue Virus (RNA extracts)



		Dengue Assay Results		
		Dengue 1	Dengue 2	Dengue 3
Sample	Dengue 1	Positive 100% (24/24)	Negative 100% (16/16)	Negative 100% (16/16)
	Dengue 2	Negative 100% (16/16)	Positive 100% (24/24)	Negative 100% (16/16)
	Dengue 3	Negative 100% (16/16)	Negative 100% (16/16)	Positive 100% (24/24)

Theranos Dengue Assays Testing with RNA Targets


Theranos primers were verified to perform well against intended pathogen Dengue targets with no cross reactivity.

Theranos Dengue assays detect Dengue 1, 2, 3 and 4 with good specificity.

Dengue Amplification/Cross Reactivity Table at 10K copies/uL

Dengue	Assay 1	Assay 2	Assay 3	Assay 4
Sample 1	Yes	No	No	No
Sample 2	No	Yes	No	No
Sample 3	No	No	Yes	No
Sample 4	No	No	No	Yes


Upper Respiratory Tests



Adenovirus




Parainfluenzavirus




Bocavirus



Respiratory syncytial virus





Influenza A virus – H1N1




Rhinovirus




Influenza A virus – H3N2



Coronavirus




Adenovirus (8 assays with DNA targets)




LoD is 12 copies/ μ l with a mean inflection time of 43.1 ± 1.4 minutes.


(7) 100 copies/ μ l target T119E1-001 amplified by P119E1-002 multiplexed with P117B1-001 (mean: 25.0 ± 0.8 min).



(8) 100 copies/ μ l target T119E2-001 amplified by P119E1-002 multiplexed with P117B1-001 (mean: 25.4 ± 0.7 min)..




Bocavirus (Assay with DNA target)



LoD is 7 copies/ μ l with a mean inflection time of 30.2 ± 1.7 minutes.


100 copies/ μ l target T126A1-001 amplified by primer set P126Ard-001 (mean: 25.4 ± 0.4 min).




Influenza A virus – H1N1 (3 assays with RNA target)

LoD is 11 copies/ μ l with a mean inflection time of 36.3 ± 2.5 minutes.


100 copies/ μ l target T122A1 RNA amplified by P122A1-001 multiplexed with P122B1 & C1 (mean: 28.4 ± 2.2 min).




100 copies/ μ l target T122B2 RNA amplified by P122B1-001 multiplexed with P122A1 & C1 (mean: 20.5 ± 0.9 min).



100 copies/ μ l target T122C1 RNA amplified by P122C1-001 multiplexed with P122A1 & B1 (mean: 20.7 ± 0.7 min).



H1N1 Assay (Specificity)





H1N1 Assay Results	
H1N1 Positive	H1N1 Negative
H1N1 sample	100% (60/60) 0% (32/32)
Cross-reactivity (24 species)	0% (0/96) 100% (96/96)

H1N1 assay with H1N1 sample

Label= P1 vs. H1N1(1pg)
Time= 32.6
Delta= 0.172
POS

Label= P1 vs. H1N1(1pg)
Time= 37.1
Delta= 0.179
POS

Label= P1 vs. H1N1(1pg)
Time= 35.0
Delta= 0.183
POS



H1N1 assay with PIV sample




Label= P1 vs. PIV_1(1pg)
Time= 0.00
Delta= 0.00
NEG

Label= P1 vs. PIV_1(1pg)
Time= 0.00
Delta= 0.00
NEG

Label= P1 vs. PIV_1(1pg)
Time= 0.00
Delta= 0.00
NEG

Label= P1 vs. no target(neg)
Time= 0.00
Delta= 0.00
NEG

Label= P1 vs. no target(neg)
Time= 0.00
Delta= 0.00
NEG



H1N1 Assay (Sensitivity)


H1N1 Assay

Performance Test	Inflection Time (min)	
	Mean	SD (1)
Negative	No Amp	nd
Positive (1 pg)		
TH1S5	19.9	0.3
Titration w/ TH1S5 (cp/ μ L)		
10,000	18.8	0.1
1,000	20.1	0.1
100	22.2	1.3
10	26.2	2.0
1	30.4	6.4


Influenza A virus – H3N2 (3 assays with RNA target)

LoD is 7 copies/ μ l with a mean inflection time of 35.1 ± 2.1 minutes.


100 copies/ μ l target T124A1 RNA amplified by P124A1-001 multiplexed with P124B1 & C1 (mean: 24.8 ± 1.3 min).



100 copies/ μ l target T124B RNA amplified by P124B1-001 multiplexed with P124A1 & C1 (mean: 22.1 ± 0.7 min).




100 copies/ μ l target T124C RNA amplified by P124C1-001 multiplexed with P124A1 & B1 (mean: 25.4 ± 2.6 min).



H3N2 Brisbane strain (inactivated virus)




H3N2 assay with H3N2 sample Negative Control Positive Control




Parainfluenzavirus (7 assays with RNA target)

LoD is 2 copies/ μ l with a mean inflection time of 42.8 ± 3.0 minutes.


(6) 100 copies/ μ l target T108A1 RNA amplified by P108A1 multiplexed with P103A1, P104B1, P105B1 & P107A1 (mean: 21.8 ± 0.4 min).



(7) 100 copies/ μ l target T108B1 RNA amplified by P108A1 multiplexed with P103A1, P104B1, P105B1 & P107A1 (mean: 28.8 ± 0.6 min).




Respiratory syncytial virus (2 assays with RNA target)




LoD is 20 copies/ μ l with a mean inflection time of 25.4 ± 2.6 minutes.



(1) 100 copies/ μ l target T109A1 RNA amplified by P109A1 multiplexed with P110A1 (mean: 17.7 ± 0.5 min).




(2) 100 copies/ μ l target T110A1 RNA amplified by P110A1 multiplexed with P109A1 (mean: 18.5 ± 0.4 min).




Rhinovirus (14 assays with RNA target)




(B) 100 copies/ μ l target T125D1 RNA amplified by P125D1-001 (mean: 21.3 ± 0.8 min).



(C) 100 copies/ μ l target T125E1 RNA amplified by P125E3-001 (mean: 17.1 ± 0.6 min).




Streptococcus pneumoniae (Assay with DNA target)




LoD is 100 copies/ μ l with a mean inflection time of 19.8 ± 1.1 minutes.

100 copies/ μ l target SPS1 DNA amplified by StrPset1 primer set (mean: 18.0 ± 1.0 min)




Penicillin-resistant Streptococcus pneumoniae (Assay with DNA target)




scienzphotolibrary

LoD is 10 copies/ μ l with a mean inflection time of 41.6 ± 9.4 minutes.

100 copies/ μ l target T120A1 DNA amplified by P120B1 primer set (mean: 28.5 ± 0.4 min)



Lower Respiratory Panel




Bordetella parapertussis



Bordetella pertussis



Enterobacter aerogenes




Escherichia coli



Staphylococcus aureus



Methicillin-resistant
Staphylococcus aureus



Moraxella catarrhalis




Mycoplasma pneumoniae




Streptococcus Group A

Bordetella parapertussis (Assay with DNA target)



LoD is 10 copies/ μ l with a mean inflection time of 25.5 ± 2.9 minutes.

100 copies/ μ l target T123B1 DNA amplified by P123C1 primer set (mean: 21.5 ± 1.5 min).




Bordetella pertussis (Assay with DNA target)




LoD is 100 copies/ μ l with a mean inflection time of 28.9 ± 2.2 minutes.

100 copies/ μ l target T123A1 DNA amplified by P123A1 primer set (mean: 25.0 ± 1.3 min).




Enterobacter aerogenes (Assay with DNA target)




LoD is 100 copies/ μ L with a mean inflection time of 27.5 ± 2.1 minutes.

100 copies/uL target T150B1 DNA amplified by P150B2 primer set (mean: 27.5 ± 2.1 min)




Escherichia coli (3 assays with DNA target)




LoD is 100 copies/ μ l with a mean inflection time of 35.0 ± 7.9 minutes.

100 copies/ μ l target T136A1 amplified by P136A1 multiplexed with P136B2 & C1 (mean: 22.3 ± 2.4 min).



100 copies/ μ l target T136B1 amplified by P136B2 multiplexed with P136A1 & C1 (mean: 18.4 ± 1.1 min).




Staphylococcus aureus (Assay with DNA target)



LoD is 10 copies/ μ L with a mean inflection time of 30.7 ± 2.1 minutes.

100 copies/ μ L target T101A1 DNA amplified by P101A1 primer set (mean: 19.2 ± 1.3 min).




Methicillin-resistant *Staphylococcus aureus* (Assay with DNA target)



LoD is 10 copies/ μ l with a mean inflection time of 30.7 ± 2.1 minutes.

100 copies/uL target T102 DNA amplified by P102B1 primer set (mean: 21.0 ± 0.9 min).




Moraxella catarrhalis (2 assays with DNA target)



LoD is 10 copies/ μ l with a mean inflection time of 29.9 ± 1.9 minutes.

100 copies/uL target T131B1 DNA amplified by P131B1 primer set (mean: 26.6 ± 0.9 min).




Mycoplasma pneumoniae (Assay with DNA target)



LoD is 5 copies/ μ L with a mean inflection time of 35.2 ± 0.5 minutes.

100 copies/ μ L target MPS1 DNA amplified by MyPset1 primer set (mean: 25.8 ± 3.1 min).




Streptococcus Group A (Assay with DNA target)



LoD is 100 copies/ μ l with a mean inflection time of 28.6 ± 1.3 minutes.

100 copies/ μ l target SPS1 DNA amplified by PStrA primer set (mean: 25.0±1.3 min)



Strep A (inactivated bacteria)




Strep A Assay


Performance Test	Mean
Negative	No Amp

Extract (cp/ μ L)	(min)
800	24.0
80	28.3
8	40.7

Respiratory Panel




Influenza A virus – H1N1 Seasonal



Influenza A virus – H1N1 Novel



Influenza A virus – H3N2




Influenza B virus



Mycobacterium tuberculosis




H7N9 Flu



H5N1 Flu



Bordetella pertussis



Methicillin-resistant
Staphylococcus aureus



Staphylococcus aureus




Adenovirus




Streptococcus Group A

Influenza A virus – H1N1 Seasonal




Limit of Detection






* Target used is the slowest of the cross targets

Inclusivity/Cross Target Detection (viral lysate eluates)



	A/New Caledonia	A/PR	A/FM	A/NWS	A/Denver	B/Lee	A/California	A/Port Chalmers	NTC
Avg	9.90	12.67	12.88	13.53	11.20	51.46	48.80	55.30	57.85
St dev	0.27	0.33	0.34	0.15	0.57	6.54	10.88	5.09	6.42



Influenza A virus – H1N1 Novel



Influenza A virus – H3N2




Limit of Detection




Lysate at ~10 ⁴ copies/ul	Average (minutes)	Standard Deviation
A/Port Chalmers/1/1973	29.75	1.79
A/Aichi/2/1968	30.51	1.03
A/Victoria/3/1975	32.30	0.70
A/Hong Kong/68	34.41	2.20
A/Brisbane/10/2007	35.39	0.61
A/Wisconsin/67/2005	35.81	1.19
A/New Caledonia/20/1999(H1N1)	63.99	5.84
A/California/07/2009 (H1N1)	68.80	7.20
B/Lee/40	75.84	8.24
NTC	70.11	11.10

Influenza B virus




Limit Of Detection



* Target used is FluB/Brisbane lysate
Copies/ul RNA, NTC-No template control

Cross Target Detection





Lysate	Avg	SD
B/Brasbie	29.42	2.61
B/R75	24.93	1.44
B/Wisconsin	17.36	0.53
A/Port Chalmers	53.93	3.62
NTC	85.06	11.96

Mycobacterium tuberculosis




TB Cross Reactivity Test




Cross reactivity test summary

	Final Copy Number	Average Inflection Time	Standard Deviation
Adenovirus	2.03×10^5	46.39	4.59
<i>Candida albicans</i>	10^5	48.53	4.96
Coronavirus	unknown (straight lysate)	49.06	8.06
<i>E. coli</i>	10^5	49.17	2.22
hgDNA	5ng	47.53	4.53
Rhinovirus	unknown (straight lysate)	48.74	4.29
Influenza A	10^5	50.55	6.20
Influenza B	10^5	49.48	7.22
<i>Pseudomonas aeruginosa</i>	10^5	47.59	3.95
<i>Staphylococcus aureus</i>	10^5	45.83	4.11
<i>Streptococcus pyogenes</i>	10^5	49.72	4.69
MTb 10^3	10^3	14.02	0.34
NTCs	-	46.28	2.34

H7N9 Flu




NA gene




HA gene

H7N9 Flu




	Final Copy Number	Average Inflection Time	Standard Deviation
Adenovirus	2.03X10 ⁵	40.25	2.36
<i>Candida albicans</i>	10 ⁵	38.09	2.54
Coronavirus	unknown (lysate 1:10)	38.39	4.35
<i>E. coli</i>	10 ⁵	37.91	2.48
hgDNA	5ng	37.96	4.01
Rhinovirus	unknown (lysate 1:10)	41.06	3.52
Influenza A	10 ⁵	39.30	3.38
Influenza B	10 ⁵	40.15	2.99
<i>Pseudomonas aeruginosa</i>	10 ⁵	37.89	1.50
<i>Staphylococcus aureus</i>	10 ⁵	37.30	3.89
<i>Streptococcus pyogenes</i>	10 ⁵	38.56	4.97
H7N9 - NA gene 10 ³	10 ³	13.59	0.61
NTCs	-	38.38	3.77



	Final Copy Number	Average Inflection Time	Standard Deviation
Adenovirus	2.03X10 ⁵	52.63	6.96
Coronavirus	unknown (straight lysate)	51.08	8.92
<i>E. coli</i>	10 ⁵	46.42	8.10
hgDNA	5ng	55.23	3.57
Rhinovirus	unknown (straight lysate)	53.27	4.96
Influenza A	10 ⁵	52.75	6.61
Influenza B	10 ⁵	51.10	8.79
<i>Pseudomonas aeruginosa</i>	10 ⁵	54.54	7.13
<i>Staphylococcus aureus</i>	10 ⁵	52.38	7.11
<i>Streptococcus pyogenes</i>	10 ⁵	56.04	4.95
H7N9 HA gene	10 ²	17.00	1.39
NTCs	-	57.87	7.65


H5N1 Flu



Tested against 3 strains of H5N1


	10^4	NTC
H5N1/Guangxi	20.09	50.76
H5N1/Hong Kong	15.56	60.03
H5N1/Xinjiang	16.41	59.15

Detection of H5N1 Influenza




Copies/ul RNA, NTC-No template control

H5N1 Cross Reactivity Test




	Copy/ μ	Average Inflection Time	Standard Deviation
Adenovirus	10^4	100.00	0.00
Candida albicans	10^5	96.70	0.00
Klebsiella pneumoniae	10^5	100.00	0.00
E. coli	10^5	98.29	4.85
hgDNA	2ng	81.42	21.11
Bordetella pertussis	10^4	98.09	5.41
A/W/S/33 (H1N1)	10^5	100.00	0.00
B/Hubei-Wujiangang/158/2009	10^5	98.07	5.45
Pseudomonas aeruginosa	10^5	100.00	0.00
Staphylococcus aureus	10^5	100.00	0.00
Streptococcus pyogenes	10^5	96.15	10.89
T166A 10^2	10^2	27.33	2.48
NTCs	-	92.32	15.35

Staphylococcus aureus



Cross-Reactivity T101 - *Staphylococcus aureus*




	Final Copy Number	Average Inflection Time	Standard Deviation
Adenovirus	2.03×10^5	66.13	8.30
<i>Candida albicans</i>	10^5	64.34	8.65
Coronavirus	unknown (lysate 1:10)	65.99	10.67
<i>E. coli</i>	10^5	66.33	5.64
hgDNA	5ng	69.36	14.09
Rhinovirus	unknown (lysate 1:10)	77.60	15.72
Influenza A	10^5	64.32	9.90
Influenza B	10^5	32.31	6.55
<i>Pseudomonas aeruginosa</i>	10^5	68.61	9.49
<i>Staphylococcus aureus</i>	10^5	9.15	0.45
<i>Streptococcus pyogenes</i>	10^5	66.61	8.30
<i>Staphylococcus aureus</i>	10^2	15.91	1.15
NTCs	-	72.31	19.56

Streptococcus Group A




Limit of Detection



Copies/ μ l DNA, NTC-No template control

Cross Target Detection (bacterial vs. viral lysate eluates)





Theranos Influenza Panel – Clinical Samples

	FluB	H1S	H3S	H1Nov	Summary	Time
Sample1	-	-	-	-	Neg	
Sample2	-	-	-	-	Neg	
Sample3	-	+	-	-	H1S	7.53
Sample4	-	+	-	-	H1S	7.31
Sample5	-	+	-	-	H1S	9.94
Sample6	-	+	-	-	H1S	9.6
Sample7	-	+	-	-	H1S	8.23
Sample8	-	-	-	-	Neg	
Sample9	-	-	-	-	Neg	
Sample10	-	-	-	-	Neg	
Sample11	-	-	-	-	Neg	
Sample12	-	-	-	-	Neg	
Sample13	-	-	-	-	Neg	
Sample14	-	-	-	-	Neg	
Sample15	-	-	-	-	Neg	
Sample16	-	-	-	-	Neg	
Sample17	-	-	-	-	Neg	
Sample18	-	-	-	-	Neg	
Sample19	-	-	-	-	Neg	
Sample20	-	-	-	-	Neg	
Sample21	-	-	-	-	Neg	
Sample22	-	-	-	-	Neg	
Sample23	-	-	-	-	Neg	
Sample24	-	-	-	-	Neg	
Sample25	+	-	-	-	FluB	11.79
Sample26	+	-	-	-	FluB	10.24
Sample27	+	-	-	-	FluB	14.71
Sample28	-	-	-	-	Neg	

BioFire	TNAA	Success Rate vs. BioFire
FluA H1N1 Pandemic	Negative	Negative
FluA H1N1 Pandemic	Negative	Negative
FluA H1N1 Pandemic	Positive	Positive
FluA H1N1 Pandemic	Positive	Positive
FluA H1N1 Seasonal	Positive	Positive
FluA H1N1 Seasonal	Positive	Positive
FluA H1N1 Seasonal	Positive	Positive
FluA H1N1 Seasonal	Negative	Negative
FluA H1N1 Pandemic	Negative	Negative
FluA H1N1 Pandemic	Negative	Negative
FluA H1N1 Seasonal	Negative	Negative
FluA H1N1 Seasonal	Negative	Negative
FluA H1N1 Seasonal	Negative	Negative
FluA H1N1 Seasonal	Negative	Negative
FluA H1N1 Seasonal	Negative	Negative
FluA H1N1 (novel)	Negative	Negative
FluA H1N1 (novel)	Negative	Negative
FluH3N2	Negative	Negative
FluB	Positive	Positive
FluB	Positive	Positive
FluB	Positive	Positive
FluB	Negative	Negative


100%

Strep B






	Avg	StDev
10^3	12.65	0.81
10^2	15.95	0.63
10	19.71	2.53
NTC	36.55	2.29

Group B Strep - Cross Reactivity Test




Panel B - STD	Final Copy Number	Average Inflection Time	Standard Deviation
<i>Candida albicans</i>	2.03×10^5	73.48	15.49
<i>Escherichia coli</i>	10^5	96.47	12.18
Herpes Simplex Virus 2	unknown (straight lysate)	100.00	0.00
<i>Enterobacter aerogenes</i>	10^5	90.34	15.13
Sng human genomic DNA	Sng	85.20	12.78
Human papillomavirus	unknown (straight lysate)	91.59	13.47
<i>Lactobacillus casei</i>	10^5	93.36	9.68
<i>Neisseria gonorrhoeae</i>	10^5	92.93	13.12
<i>Pseudomonas aeruginosa</i>	10^5	100.00	0.00
<i>Staphylococcus aureus</i> MSSA (DmecA)	10^5	96.56	6.53
<i>Streptococcus agalactiae</i>	10^5	13.12	0.50
<i>Strep agalactiae</i>	10^2	24.66	1.56
No Template Control	-	100.00	0.00

Testing Potential Interfering Substances




Potential carry over substances through sample prep does not inhibit the assays.

Testing Potential Interfering Substances - Strep A Assay



STD Panel





Hepatitis B Virus




Human immunodeficiency virus

Hepatitis B Virus (3 assays with DNA targets)


100 copies/uL target T137A1 DNA amplified by P137A1 multiplexed with P137B1 & C1 (mean: 24.8 ± 1.2 min)
LoD is 100 copies/uL with a mean inflection time of 32.8 ± 1.4 minutes.



100 copies/uL target T137B1 DNA amplified by P137B1 multiplexed with P137A1 & C1 (mean: 22.0 ± 1.0 min)




100 copies/uL target T137C1 DNA amplified by P137C1 multiplexed with P137A1 & B1 (mean: 32.8 ± 1.4 min)




Human immunodeficiency virus (3 assays with RNA targets)


1 000 copies/ μ l target T140Δ1 RNA amplified by P140Δ1 primer set (mean= 29.8+1.2 min)



10,000 copies/ml T440D4 RNA was used by D440D4 animals at 1/1000 dilution. 07/10/7 min.




1 000 copies/ul target T440A4 RNA quantified by T440A4 primers set (mean: 27 210.7 min)




appendix


Clinical Correlation - Serum and Amniotic samples




CK-MB



C-reactive protein



Estradiol



Estradiol - Precision

Analyzer	Signal (RLU)	Difference from Mean	Conc. pg/mL
	Mean RLU		
1	9287	-5	74
2	9523	-2	71
3	9991	2	66
4	10160	4	64
5	9397	-4	72
6	10383	7	62
7	10275	5	63
8	10351	6	62
9	8432	-14	84
10	8334	-15	86
11	9041	-7	77
12	9457	-3	72
13	10687	10	59
14	10984	13	57
15	10704	10	59
16	10823	11	58
17	10125	4	65
18	10222	5	64
19	10168	4	64
20	10452	7	61
21	8832	-9	79
22	8365	-14	85
23	8425	-14	84
24	9562	-2	71

Inter-Analyzer Concentration CVs

Mean Conc. pg/mL	StD ev	CV%	% Recovery
69	9	13	92

Inter-Analyzer Signal CVs

Mean RLU	StDev	CV%
9749	835	9


Estradiol – Precision

QC Levels for 3 Day Precision and Accuracy


Nominal [Estradiol] pg/mL	Cartridg e	Recovered [Estradiol] pg/mL					
		Day 1	Day 2	Day 3	Mean Conc.	CV %	% Recovery
478.50	1	522	481	448	455	8	95
	2	402	426	485			
	3	440	444	444			
92.63	1	114	104	72	92	14	99
	2	104	91	85			
	3	90	81	87			
45.38	1	36	37	37	41	11	90
	2	40	38	44			
	3	44	48	44			

Unconjugated Estriol

Theranos vs Siemens




Theranos vs DRG




Human IgG


Clinical samples correlation




Free T3



Free Testosterone




Folic Acid




Human IgG

Clinical samples correlation




Magnesium


Magnesium clinical correlation




Mumps antibody




Total Protein



Prealbumin




Free Thyroxine (fT4)




Troponin I (cTNI)

cTNI Clinical Correlation: Theranos to Commercial Methods reported value
(Beckmann Access and Vista Dimension)
(Total 103 samples tested)




Theranos cTNI Assay: Clinical Sample Correlation to Individual Commercial Methods


Li-Hep Plasma samples correlation:
Theranos to Beckman Access (N=29)



Li-Hep Plasma samples correlation:
Theranos to Vista Dimension (N=27)




cTNI Clinical Correlation: Theranos to Siemens
Immulite
(Total 103 samples tested, N=79 in plot with
CLIA value >0.2ng/ml)




Troponin T


Clinical sample correlation: Theranos vs. Roche cTNT assay



Total Valproic Acid (VA)



Vancomycin



25-OH Vitamin D

