

From: Surekha Gangakhedkar <surekhag@theranos.com>
Sent: Monday, May 10, 2010 11:11 AM
To: Victoria Sung <VSung@celgene.com>
Cc: Gary Frenzel <gfrenzel@theranos.com>
Subject: RE: Thank you
Attach: CONFIDENTIAL - Theranos Celgene Assay Update May 7 2010.pdf

Hi Vicki,

We believe that the Friday meeting was very productive. Here is a copy of the Presentation. Please note that all the information is Theranos Confidential. We are looking forward to hearing back from you on the items noted below.

Best regards,
 Surekha

Surekha Gangakhedkar
 Team Manager, Assay System
 Theranos Inc.
 3200 Hillview Avenue
 Palo Alto, CA 94304
 Phone:650-838-9292
 Fax: 650-838-9165

PRIVILEGED AND CONFIDENTIAL COMMUNICATION IMPORTANT – This electronic transmission, and any files transmitted with it are confidential and/or legally privileged information. This information is intended solely for the use of the individual or entity to which they are addressed. Any disclosure, retransmission, reproduction, dissemination or other use of the contents of this information by persons or entities other than the intended recipient is strictly prohibited. If you have received this email in error, please contact us immediately and delete all copies. Please note that any views or opinions presented in this email are solely those of the author and do not necessarily represent those of Theranos, Inc. Finally, before opening or using attachments the recipient should check this email and any attachments for the presence of viruses. Theranos, Inc. accepts no liability for any damage caused by any virus transmitted by this email. Our sole responsibility is limited to resupplying any affected attachments.

Theranos, Inc., 3200 Hillview Ave, Palo Alto, CA 94034
 650-838-9292 www.theranos.com

=====

From: Victoria Sung [mailto:VSung@celgene.com]
Sent: Friday, May 07, 2010 5:31 PM
To: Gary Frenzel; Surekha Gangakhedkar; Daniel Young
Cc: Elizabeth Holmes
Subject: Thank you

Thank you all for your time this morning; we felt that our meeting was very productive and informative! We will be following up on a number of items we discussed earlier today:

- 1) Finalize the complete list of analytes to be assayed (Parts 1 and 2)
 - Testosterone vs. DHT vs. DHEA?
 - Estrogen vs. estradiol (Theranos only developing estradiol assay)?
 - FGF23: intact vs. cleaved?
 - EPO: do we want to exclude recognition of recombinant EPO (only detect endogenous EPO)?
 - IGF: current assay detects only IGF bound to IGF-BP3 (not 1/2)?
- 2) Part 1: planning to use Theranos technology to assay PK, ADA **plus** Hgb, Vitamin D, FSH, sex steroids, PTH (validate these PD markers by comparing vs. ACM-generated data)
 - Can we amend protocol to reflect that PD markers may be measured in Part 1 (i.e. extra blood draw in a pediatric tube for Theranos assays)
 - Will part 1 sites be ok with loading multiple machines to look at PD markers (in addition to PK and ADA)?
- 3) For Theranos: testosterone to develop the assays (they are still waiting for a license to purchase steroids), Epogen or

Procrit to test if their assay recognizes recombinant protein, ACE-011 plus mouse/goat serum for ADA assay development (if possible, serum from monkey treated with ACE-011), the opportunity to "validate" their PD assay controls at ACM (Sharianne to arrange for work order with ACM).

4) We will also find out who should be on the user list for access to the web portal (blinded or unblinded) and also a list of patient ID# to pre-program into the machines for Part 1.

5) Amend 3-way NDA (Celgene/Theranos/Ron Bowsher) to expand scope of discussion to validation of PD biomarkers and data.

If there are other questions not included above, please feel free to e-mail me. In the meantime, would you please send me your presentation from this morning and also let me know when I should expect the validation reports to start trickling in? Do you think that it would be valuable to set a teleconference for later this month to address the above issues and anything else that develops between now and then?

Have a good weekend!
Vicki

THIS ELECTRONIC MAIL MESSAGE AND ANY ATTACHMENT IS
CONFIDENTIAL AND MAY CONTAIN LEGALLY PRIVILEGED
INFORMATION INTENDED ONLY FOR THE USE OF THE INDIVIDUAL
OR INDIVIDUALS NAMED ABOVE.
If the reader is not the intended recipient, or the
employee or agent responsible to deliver it to the
intended recipient, you are hereby notified that any
dissemination, distribution or copying of this
communication is strictly prohibited. If you have
received this communication in error, please reply to the
sender to notify us of the error and delete the original
message. Thank You.



Celgene panel: Assay development update

May 7th, 2010

Theranos, Inc.

This presentation and its contents are Theranos proprietary and confidential



Targeted Celgene Panels

Panel for part 1

	Analyte
1	ACE-011 PK
2	anti-ACE-011 antibody
3	Hgb
4	EPO
5	PTH
6	Vit D
7	FSH
8	LH
9	Estradiol
10	DHT
11	Testosterone

Panel for part 2

	Analyte		Analyte
1	ACE-011 PK	12	Hepcidin
2	anti-ACE-011 antibody	13	BSAP
3	Hgb	14	CTX
4	EPO	15	NTX
5	PTH	16	Trap5b
6	Vit D	17	Osteocalcin
7	FSH	18	P1NP
8	LH	19	P1CP
9	Estradiol	20	FGF-23
10	DHT	21	IGF-1
11	Testosterone		



theranos
redefining healthcare

Assay Development Process


	Experiment	Information desired
1	MTP screen	Summary information of antibody screens, 4 point curves
2	Conjugation Type: SH or NH2	Compare SH vs. NH2 conjugations (For assays with limited antibodies to screen)
3	Cross reactivity	For 4 best Ab. Pairs: Test cross reactivity with related analytes, 6 point standard curves
4	Interference	For 4 best Ab. Pairs: Test interference with related analytes, 6 point standard curves
5	Epitope map	For 4 best Ab. Pairs
6	Theranos screen	For 3 best Ab. Pairs: Summary information for 6 point curves in A.B, whole blood and plasma
7	Training set	For 1 Ab. Pair correlation of N=10 clinical samples spanning the assay range range
	<i>Antibody Pair finalized</i>	<i>Based on the best dose response and % recovery of above experiments</i>
8	Plasma Screen	Screen normal plasma to see ranges
9	Whole Blood Screen	Screen normal blood to see ranges.
10	Capture titration	Show table with selected level and higher/lower levels
11	Detection titration	Show table with selected level and higher/lower levels
12	Detection Stabilizer	Biostab vs. Stabilzyme
13	Effect of Buffer	Test various blocking buffers during coating, Assay buffer compositions
14	Effect of Incubation time	Compare different incubation times
15	Edison protocol optimization	Different dilution effects, impact of PSW
16	% CV	Mid range analyte on 24 instruments
17	Precision tests	3 lot inter and intra assay results with standard curve, 3 cartridges per lot
18	Calibrator comparison	Run our controls on reference kit and vice versa, calibrate with WHO standards
19	Dilution linearity	Serial dilution of high samples in buffer and sera
20	Whole blood spike recovery	Spike into blood with known (low) endogenous level at 8 point curve
21	Hematocrit effect	Plasma from spiked blood tested, plot W.B vs. plasma
22	Plasma spike recovery	Spike into plasma with known (low) endogenous level at 8 point curve
23	LLOQ/ULOQ	Use 2x above ULOQ and below LLOQ to define assay limits
24	Selectivity	Spike into W.B at 3 levels over endogenous to calculate recovery
25	Stability	CAB, DAB stability, Reference tips in some cases, with a 4-point standard curve
26	Matrix effects	Spike recovery in hemolysed or lipemic samples for 6 spike levels
27	Extended range	Test linearity and dose response of standard curve at 4x of ULOQ
28	Clinical samples	N=20, spanning the range to show correlation with kit




ACE-011 update

- Acceleron-QPS assay range: 8 – 400 ng/mL
- Competitive ELISA
- Theranos assay range: 40 – 4000 ng/mL
- Both Sandwich and Competitive ELISA formats were evaluated
- Competitive assay was selected for Theranos assay as well
- Assay reagents for Theranos assay similar to used by QPS except the label is Alkaline Phosphatase.

Representation of QPS ACE-011 assay



Representation of Theranos ACE-011 assay





ACE-011: Precision test data

- Pooled serum 8- point standard curve
- 5 QC levels that were within the range of the assay
- QC levels generated by spiking ACE-011 into pooled serum
- 3 cartridges per day, duplicate data points for each cartridge. 2 days total.
- Back calculated concentrations of the QC levels obtained from standard curve.



ACE-011 Precision test: Standard curve data

AVERAGE CONC. FOR EACH CARTRIDGE										
Conc. ng/mL	Day 1			Day 2			Average	Stdev	%CV	% Recovery
	1	2	3	4	5	6				
8000	8066	8484	8545	7551	7300	8171	8019	501	6	100
4000	3973	4059	4299	3754	3624	3833	3924	240	6	98
1200	1445	1391	1351	1258	1131	1054	1272	154	12	106
400	411	365	447	331	335	325	369	50	13	92
80	71	80	71	64	61	69	69	6	9	87
40	52	45	51	51	53	49	50	3	5	125
20	16	17	22	21	18	17	19	3	14	93
0										




ACE-011 Precision test: QC levels data

		AVERAGE CONC. FOR EACH CARTRIDGE									
	Conc. ng/mL	Day 1			Day 2						
		1	2	3	4	5	6	Average	Stdev	%CV	% Recovery
ULOQ	4000	4639	4637	4485	3668	3599	3598	4104	532	13	103
QCH	2000	2247	2383	2580	2498	2133	2265	2351	168	7	118
QCM	600	580	695	628	585	697	662	641	52	8	107
QCL	160	151	146	154	166	147	145	152	8	5	95
LLOQ	40	52	45	51	51	53	49	50	3	5	125



ACE-011


Comparison of whole blood spike of N=6 samples to QPS system





ACE-011

Comparison of plasma from spiked whole blood of N=6 samples to QPS system





2. Bone specific Alkaline Phosphatase

- **Bone AP Clinical Range:**

Normal Female (≥ 25)=11.6 – 42.7 U/L

Normal Female (≥ 25)=15 – 41.3 U/L

Paget’s Disease of Bone: Median 199.6 U/L

Osteomalacia: Mean 61.7 U/L

Osteoporosis (no anti-resorptives s): 6 – 25.9 U/L

Osteoporosis (On anti-resorptives): 4 – 17.9 U/L

- **Theranos Assay Range:** 200 – 3.125 U/L

- **Theranos Reference Assay 1 :** MicroVue BAP Kit. Vendor: Quidel, Range: 140 – 2U/L

- **Theranos Reference Assay 2:** Ostase BAP Kit. Vendor: IDS, Range: 90 – 7ug/L

Please note that the Theranos assay and the reference assays are direct assays that use capture only, no detection conjugate. All three assays use the activity of the analyte itself.

	Antibody	SH- Conj	NH2- Conj
Capture Ab	#1		
	#2		
	#3		
	#4		
	#5		
	#6		
	#7		
	#8		
	#9	N/A	
	#10	N/A	

Number of Capture antibodies tested: 10

Number of Detection antibodies tested: 0

Total Number of conjugates tested: 18

- No Response
- Poor response but potential
- Modulation but background or other problem
- Modulation, good candidate pair
- Modulation, good candidate pair




BSAP: Interference Testing

For interference testing, the stated level of secondary substances was spiked into a six point standard bone AP curve and the resulting solutions tested against a standard control curve.

Substance Tested	Level Tested	Avg %Recovery of BAP	Interference?
Alendronate	50ug/mL	85	No
Estradiol	100ug/mL	92	No
Acetaminophen	350ug/mL	88	No
TRAP 5b	31.8 U/L	94	No
Ace-011	24,000ng/mL	87	No
Liver Alkaline Phosphatase	400 U/L	93	Yes at BAP Levels 3.125 U/L and below
Intestinal Alkaline Phosphatase	1000 U/L	96	Yes at 12.5 U/L and below



BSAP: Dilution Linearity: Serum and Buffer



Summary of Results for Dilution Linearity Diluting Clinical Serum #26

Dilution Matrix	Dilution	Calculated [BAP] U/L	% Recovery
Assay Buffer	None	65.33	[100]
	2X	25.86	93
	4X	13.72	99
	8X	6.96	101
	16X	2.52	73
	32X	1.33	77
Serum	None	63.54	[100]
	2X	39.47	117
	4X	18.63	99
	8X	13.56	119
	16X	8.77	114
	32X	4.97	86


Theranos Confidential





theranos
redefining healthcare

BSAP: Matrix Effects

Matrix=>	Hemolyzed Serum	Lipemic Serum
Activity (U/L) Spiked	%Recovery	%Recovery
92.500	92	117
61.667	96	101
41.111	92	118
27.407	111	67
18.272	103	90
12.181	88	102
Avg% Recovery	97	99



BSAP: Capture antibody Titration



Capture antibody Titration: Assay Buffer Standards

Capture Concentration (ug/mL)	2.5	5	10	20
Std 1/6	1703	2399	2006	2083
Std 5/6	10.18	14.27	11.20	11.16
Average Ratio	4.94	5.83	5.18	5.26
Avg % Signal CV	11.7	12.6	12.8	10.1
Avg % Conc CV	9.32	13.38	10.94	7.96
LOD in Sample (U/L)	0.53	0.49	0.35	0.30
Avg% Recovery	114	95	101	112

3. EPO

- **Clinical sample ranges:**
- Normal subjects: EPO varies from 3 -16 mU/mL.
- Subjects with anemia: Aplastic anemia, hemolytic anemia and anemia due to iron deficiency all result in serum EPO elevation. The normal kidney's response to anemia is that when Hct levels fall <20% EPO levels increase normal, at times as high 50-500 mU/mL. There is an inverse relationship between the hematocrit level and EPO.
- RA, inflammation patients: 20-200 mU/mL
- Patients with CKD/ESRD: Anemia is a common side effect in these patients. In these patients the kidney does not have a normal response to anemia. Hence the levels are close or lower than normal.
- Leukemia patients: Anemia is a common condition in these patients as a result of chemotherapy. High levels of EPO are seen in patients after successive rounds of chemotherapy (~200 mU/mL).
- **Theranos assay range:**
- Spiked plasma/serum calibrators: 12 point curve: 250-1.25 mU/mL.
- Assay will be calibrated against NIBSC International standard 87/684 (recombinant hEPO).
- **Reference Commercial ELISA: R&D Quantikine IVD human EPO ELISA**
Reported sensitivity: 0.6 mU/mL , time: Benchtop protocol 4h 45' Sample volume per well 100 μ L



EPO: Antibody screen summary

Capture Ab	Detection Ab																	
	#1	#2	#3	#4	#5	#6	#7	#8	#9	#10	#11	#12	#13	#14	#15	#16	#17	#18
#1																		
#2																		
#3																		
#4																		
#5																		
#6																		
#7																		
#8																		
#9																		
#10																		
#11																		
#12																		
#13																		
#14																		
#15																		
#16																		
#17																		
#18																		



theranos
redefining healthcare

EPO: Antibody pair cross reactivity

Ab pair 1

Nominal mU/mL	Cross reactivity Test analyte		1250-4.88 pg/mL		5000-5 pg/mL		1000-1 pg/mL		750-0.5 pg/ml		1000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL	
	Calc. (mU/mL)	%recovery	Calc.(mU/mL)	% Cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc.(mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity
250	245	98	1.37	1	1.05	0	0.88	0	0.38	0	2.40	1	2.09	1	1.72	1	0.87	1
100	105	105	1.14	1	1.09	1	0.92	1	0.83	1	2.21	2	1.98	2	1.50	1	1.52	1
50	50	101	1.75	3	1.48	3	0.89	2	0.97	2	3.48	7	2.38	5	1.95	4	1.62	2
20	19	95	1.93	19	1.45	7	1.04	5	1.03	5	2.84	13	2.36	12	2.12	11	1.78	2
5	5	102	1.98	40	1.62	32	1.30	26	1.20	24	4.03	81	2.85	53	2.45	49	1.88	2

Ab pair 2

Nominal mU/mL	Cross reactivity Test analyte		1250-4.88 pg/mL		5000-5 pg/mL		1000-1 pg/mL		750-0.5 pg/ml		1000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL	
	Calc. (mU/mL)	%recovery	Calc.(mU/mL)	% Cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc.(mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity
250	245	98	0.00	0.00	0.00	0.00	0.02	0.01	0.17	0.07	0.89	0.89	0.08	0.03	0.16	0.06	6.13	2.45
100	96	96	0.00	0.00	0.05	0.05	0.18	0.18	2.11	2.11	0.36	0.38	0.12	0.12	1.43	1.43	2.80	2.80
50	52	104	0.00	0.00	0.00	0.01	0.01	0.01	0.07	0.13	0.01	0.01	0.01	0.02	0.06	0.12	1.32	2.85
20	18	92	0.00	0.00	0.01	0.05	0.11	0.53	2.04	10.22	0.92	1.00	0.01	0.03	0.04	0.21	2.03	10.16
5	5	101	0.00	0.03	0.01	0.21	0.08	1.52	0.18	3.51	0.00	0.00	0.00	0.00	0.00	0.03	0.39	7.76



Ab pair 3

Nominal mU/mL	Cross reactivity Test analyte		1250-4.88 pg/mL		5000-5 pg/mL		1000-1 pg/mL		750-0.5 pg/ml		1000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL	
	Calc. (mU/mL)	%recovery	Calc.(mU/mL)	% Cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc.(mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity
250	250	100	8	3	1	1	1	0	0	0	7	7	6	2	4	2	1	0
100	101	101	6	6	2	2	1	1	1	1	6	6	5	5	3	3	3	3
50	49	98	3	7	3	6	1	2	1	2	13	14	7	15	5	10	4	7
20	20	102	3	17	3	15	1	7	1	7	9	9	7	36	6	30	4	22
5	5	99	4	84	4	72	2	46	2	39	16	16	9	176	8	154	4	78

Ab pair 4

Nominal mU/mL	Cross reactivity Test analyte		1250-4.88 pg/mL		5000-5 pg/mL		1000-1 pg/mL		750-0.5 pg/ml		1000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL		6000-5 pg/mL	
	Calc. (mU/mL)	%recovery	Calc.(mU/mL)	% Cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity	Calc.(mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-re	Calc. (mU/mL)	%cross-reactivity	Calc. (mU/mL)	%cross-reactivity
250	245	98	1.84	1	1.74	1	1.97	1	2.19	1	1.47	1	1.67	1	1.68	1	2.42	2
100	97	97	1.60	2	1.93	2	2.80	3	2.21	2	1.70	2	1.53	2	1.84	2	1.88	2
50	58	116	1.94	4	2.43	5	1.92	4	1.90	4	1.80	3	1.74	3	1.54	3	1.56	1
20	17.1	85	1.54	8	1.44	7	2.41	12	2.25	11	1.77	9	1.52	8	1.83	9	1.89	2
5	5.3	106	1.43	29	1.86	37	2.28	46	2.01	40	1.45	29	1.51	30	1.97	39	1.90	2

EPO: Assay Optimization



4. Estradiol



- Males: 8 - 43 pg/ml
- Females:
 - Follicular Phase 13 - 166 pg/ml
 - Ovulation Phase 86 - 498 pg/ml
 - Luteal Phase 45 - 211 pg/ml
 - Postmenopause: <5.0 - 54 pg/ml
 - Pregnant 1st trimester 38 - 3175 pg/ml
 - Pregnant 2nd trimester: 678 - 16633 pg/ml
 - Pregnant 3rd trimester: 43 - 33781 pg/ml
- Up to 99% of estradiol in the blood is bound to either SHBG or human serum albumin (HSA).
- 2 approaches are generally used to measure estradiol in light of these challenges:
 1. Displacement of estradiol from binding proteins by displacement agents
 2. Utilization of an anti-estradiol antibody with a sufficiently high affinity and/or an epitope that allows the antibody to bind to SHBG-bound and HSA-bound estradiol




Estradiol

- 63 commercially available anti-estradiol antibodies were screened: 32 monoclonal and 2 goat polyclonal, 7 sheep polyclonal, and 22 rabbit polyclonal antibodies
- Spiked serum calibrators were used for screening
- Displacement agents including Danazol and a commercial buffer containing DHT failed to improve S/B for the low-response antibodies
- Only 3 of the antibodies yielded > 2 Signal/BG at 2000 pg/mL in a serum matrix
- The 3 final candidates were tested in the following matrixes to determine matrix effects:
 - Spiked whole blood
 - Plasma derived from spiked whole blood
 - Pooled plasma, directly spiked
 - Pooled serum, directly spiked
 - Clinical serum samples, concentration previously determined via 2 reference methods: Alpco and Invitrogen

Estradiol Matrix Screening Results



- Cab 61 showed the most parallel slopes for all matrixes tested
- Cab 63 was eliminated due to major matrix effects
- Cab 61 and Cab 62 were chosen for further testing





Estradiol Clinical Correlation Screen

- Candidates Cab 61 and Cab 62 were tested for clinical correlation with a reference method
- 5 clinical samples from across the assay range were chosen
- Based on the low matrix effects, availability, and good clinical correlation, Capture Antibody #61 was chosen

Sample #	Sample Type	Alpco Result pg/mL	Theranos Results pg/mL	
			Cab 61	Cab 62
58	Pregnant Week 22	1455	956	1000
50	Pregnant Week 12	681	442	575
16	Menstrual Day 16	208	107	118
7	Menstrual Day 7	52	38	41
80	Post-Menopausal	29	22	4

Continued next slide

Theranos Confidential

24



theranos
redefining healthcare

Checkerboard Titration for Estradiol

- A checkerboard titration was performed with spiked serum calibrators to determine dilutions of CAb and Estradiol conjugate for initial tests
- Cab @ 1:25,000 and Estradiol-AP @ 1:50,000 was determined as optimal

CAb	Estradiol-AP	Estradiol] pg/l	Mean RLU	StDev	CV%	S/B
1:25,000	1:10,000	2031.0	15545	1467	9	9.4
		231.0	69654	9649	14	2.1
		31.0	128544	8766	7	1.1
		6.0	146240	10812	7	
1:50,000	1:10,000	2031.0	6352	1527	24	9.4
		231.0	31145	7145	23	1.9
		31.0	54489	7835	14	1.1
		6.0	59443	7832	13	
1:150,000	1:10,000	2031.0	2503	224	9	9.8
		231.0	10092	626	6	2.4
		31.0	20039	2833	14	1.2
		6.0	24545	1955	8	
1:25,000	1:25,000	2031.0	6017	557	9	13.8
		231.0	30413	3221	11	2.7
		31.0	63891	8434	13	1.3
		6.0	83247	5881	7	
1:50,000	1:25,000	2031.0	3599	76	2	12.5
		231.0	16067	705	4	2.8
		31.0	35674	136	0	1.3
		6.0	45160	1240	3	
1:150,000	1:25,000	2031.0	1331	29	2	12.2
		231.0	5328	296	6	3.1
		31.0	12597	106	1	1.3
		6.0	16308	623	4	
1:25,000	1:50,000	2031.0	4228	699	17	15.6
		231.0	21350	1727	8	3.1
		31.0	45574	4210	9	1.4
		6.0	65896	11094	17	
1:50,000	1:50,000	2031.0	2319	540	23	13.9
		231.0	11217	2707	24	2.9
		31.0	24905	4235	17	1.3
		6.0	32257	4268	13	
1:150,000	1:50,000	2031.0	1046	90	9	11.8
		231.0	3889	419	11	3.2
		31.0	9465	1261	13	1.3
		6.0	12372	919	7	
1:10,000	1:100,000	2031.0	5077	829	16	10.2
		231.0	20358	15574	77	2.5
		31.0	47242	706	1	1.1
		6.0	51762	800	2	
1:25,000	1:100,000	2031.0	1795	106	6	13.2
		231.0	8208	69	1	2.9
		31.0	20912	2263	11	1.1
		6.0	23694	1287	5	
1:50,000	1:100,000	2031.0	1093	155	14	15.0
		231.0	4607	313	7	3.6
		31.0	11922	881	7	1.4
		6.0	16431	254	2	




5. FSH

- **Normal ranges**
- **Females:**
 - Follicular phase 3.5 - 12.5 mIU/ml
 - Ovulation phase 4.7 - 21.5 mIU/ml
 - Luteal phase 1.7 - 7.7 mIU/ml
 - Postmenopause 25.8 - 134.8 mIU/ml
- **Males** 1.5 - 12.4 mIU/ml
- Theranos assay ranges: 200mIU/ml to 0.5mIU/ml
- Reference assay: Genway Biotech FSH Elisa Kit



Capture Ab	Detection Ab																					
	1	2	3	4	5	6	7	8	9	10	11	12	14	15	16	17	18	19	20	21	22	
1																						
2																						
3																						
4																						
5																						
6																						
7																						
8																						
9																						
10																						
11																						
12																						
13																						
14																						
15																						
16																						
17																						
18																						
19																						
20																						
21																						
22																						

FSH: Analyte verification





FSH: Stability

Capture Surface





Detection antibody Stability





FSH: Hematocrit effect






FSH: Precision for 3 reagent lots

Concentration (mIU/ml)	Cartridge	RLU			Mean	StDev	Intra %CV	Calc. Concentration			Total Concentration CVs			
		Lot 1	Lot 2	Lot 3				Lot 1	Lot 2	Lot 3	Mean	StDev	CV %	% Recovery
200	1	303290 314890	350590 365651	369117 316310	334859	33100	10	198	201	200	200	1	0.6	100
	2	356880 346391	300340 341572	313569 344499				201	199	199				
	3	380078 356791	278943 349508	373281 265771				201	198	199				
40	1	38368 39616	37484 39126	33409 34447	36169	3232	9	44	43	37	40	4	9.7	101
	2	36379 38866	37857 39458	26766 33057				42	44	32				
	3	36100 39123	37456 35530	34489 33512				42	41	37				
8	1	10458 11078	9826 10172	9901 8604	10227	957	9	8	8	7	8	1	11.4	98
	2	11941 11241	9717 9419	9613 9223				9	7	7				
	3	10324 12001	8842 10236	10697 10791				9	7	8				
2	1	3892 3797	3878 3257	4018 3913	3748	300	8	2.1	1.9	2.2	2	0	7.8	102
	2	3093 3587	3626 3776	3847 3545				1.7	2.0	2.0				
	3	3642 4007	3804 3539	4458 3783				2.1	2.0	2.3				
0.5	1	1239 1022	1088 1030	1010 1142	1114	85	8	0.5	0.5	0.5	0	0	5.0	100
	2	1178 1197	1098 941	1160 1180				0.5	0.5	0.5				
	3	1067 1247	1109 1051	1098 1199				0.5	0.5	0.5				
0	1	516 510	583 596	543 555	551	36	6	0	0	0	0	0	3.5	N/A
	2	631 522	586 600	515 505				0	0	0				
	3	548 538	533 569	532 544				0	0	0				



FSH: Training set of clinical samples





6. FGF-23



- Clinical range: 20 -1400 pg/mL
- Reference assays
 - Millipore ELISA, 10 - 2400 pg/mL, measures native FGF-23
 - Kainos ELISA, 10 - 800 pg/mL, Full length FGF-23
- Antibody screening data

	Dab1	Dab2	Dab3	Dab4	Dab5	Dab6	Dab7	Dab8	Dab9	Dab11	Dab12	Dab13	Dab14	Dab15	Dab16
Cab1	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab2	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab3	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab4	Red	Red	Red	Red	Red	Green	Red	Red	Red	Red	Yellow	Red	Red	Red	Red
Cab5	Red	Red	Red	Red	Red	Yellow	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab6	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab7	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab8	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab9	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab10	Red	Red	Red	Red	Red	Yellow	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab13	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab14	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab15	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
Cab16	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red


Good dose response
 Moderate dose responses
 Poor modulation or no modulation

FGF-23: Calibrator Qualification

- Calibrators used for Theranos assay development are qualified against standards from Reference ELISA kit
- When available, analyte is qualified against WHO reference standards



FGF-23: Effect of Matrix





7. Hemoglobin

Normal & Clinical Ranges

Male – 140-150 mg/ml normal

130-175 mg/ml clinical

Female – 135-145 mg/ml normal

120-160 mg/ml clinical

Theranos Assay Range

50 mg/ml – 250 mg/ml

Reference ELISA

Arbor Assays DetectX Hb Kit

Range: 0.25-16 g/dL (2.5-160 mg/ml)

Antibody Screen Summary							
Capture	Detection						
	1D	2D	3D	4D	5D	6D	7D
1C	Red	Yellow	Red	Red	Red	Yellow	Red
2C	Yellow	Yellow	Red	Red	Yellow	Yellow	Yellow
3C	Yellow	Yellow	Red	Red	Green	Yellow	Yellow
4C	Red	Red	Red	Red	Yellow	Red	Red
5C	Yellow	Yellow	Red	Red	Yellow	Yellow	Yellow
6C	Yellow	Yellow	Red	Red	Green	Yellow	Green
7C	Green	Yellow	Red	Red	Green	Yellow	Red


	Poor, no modulation
	Satisfactory pair, will do if need be; may need more signal, or high background; modulation okay
	Good candidate pair; good signal, good modulation; good S/B



Hemoglobin Assay Data: D-Ab Titration

50 ng/ml D-Ab					
[Hb], ug/ml	Tip 1	Tip 2	MeanValue	StDev	CV
10	423,561	479,886	451,723	39,828	9%
8	396,188	437,885	416,941	29,350	7%
6	360,913	372,323	366,618	8,069	2%
4	274,037	333,798	303,918	42,258	14%
2	198,480	183,279	190,880	10,748	6%
0	30,862	28,934	29,898	1,363	5%

20 ng/ml D-Ab					
[Hb], ug/ml	Tip 1	Tip 2	MeanValue	StDev	CV
10	185,739	205,882	195,811	14,243	7%
8	178,181	192,013	185,097	9,780	5%
6	149,992	166,412	158,202	11,511	7%
4	124,511	134,000	129,256	6,710	5%
2	71,124	63,480	67,292	5,419	8%
0	12,127	12,750	12,439	440	4%






Hemoglobin Assay Data: Precision

Instrument	Tip 1	Tip 2	Mean Value	StDev	CV
1	459,592	468,692	464,142	6,435	1%
2	732,104	678,129	713,045	38,167	5%
3	700,275	753,520	726,897	37,650	5%
4	618,048	692,613	655,331	52,726	8%
5	619,372	665,056	642,514	31,879	5%
6	604,632	655,197	629,915	35,755	6%
7	765,795	761,147	763,471	3,286	0%
8	729,940	725,110	727,525	3,416	0%
9	748,514	853,418	800,966	74,179	9%
10	757,282	748,643	752,963	6,108	1%
11	735,558	708,053	720,806	20,863	3%
12	657,125	710,617	683,871	37,824	6%
13	785,322	623,868	704,595	114,165	16%
14	793,571	678,926	736,248	91,066	11%
15	676,690	698,957	687,823	15,745	2%
16	810,560	769,766	800,163	14,703	2%
17	732,005	540,400	636,202	135,485	21%
18	713,245	669,382	691,314	31,016	4%
19	519,450	411,689	465,569	76,199	16%
20	854,783	686,870	670,827	22,689	3%
21	704,591	574,090	639,340	92,278	14%
22	549,361	770,325	659,843	156,245	24%
23	702,559	716,568	709,564	9,906	1%
24	731,676	673,044	702,360	41,459	6%

AVERAGES 683,970 47,469 7%






8. Insulin like growth factor (IGF-1)

- Clinical range for normals ranges from 10 -1000 ng/mL
- Theranos range is 12 -1200 ng/mL
- Approximately 98% of IGF-1 is always bound to one of 6 binding proteins (IGF-BP). IGFBP-3, the most abundant protein, accounts for 80% of all IGF binding. IGF-1 binds to IGFBP-3 in a 1:1 molar ratio.
- Sandwich Elisa established with Capture Antibody specific to IGF-1 and the Detector antibody specific to the IGF Binding Protein 3 so that the complex can be accurately measured in sample.
 - Eliminates the need for a purification/separation step which all commercial Elisa Kits require.




IGF-1: Whole Blood and Plasma Screens



Sample #	Whole Blood (ng/mL)	Plasma (ng/mL)
1	216	214
2	161	152
3	152	224
4	108	162
5	72	102
6	106	176
7	211	69
8	147	181
9	116	236
10	60	117



IGF-1: Precision for 3 reagent lots



Concentration CV's

Std	Conc. (ng/mL)	Lot 1	Lot 2	Lot 3	Avg
1	1800	6%	6%	20%	10%
2	1200	8%	17%	10%	12%
3	600	4%	12%	5%	7%
4	300	11%	8%	8%	9%
5	150	8%	13%	12%	11%
6	75	18%	19%	4%	14%
7	38	11%	7%	29%	16%
8	19	4%	5%	29%	13%
9	10	9%	7%	19%	12%
10	5	11%	5%	23%	13%
11	2.5	9%	2%	25%	12%
12	0	11%	2%	28%	14%




IGF-1: 24 Instrument CV's

CV 24 Instruments								
Tray #	Conc. (ng/mL)	Signal MeanValue	Signal Std.Dev.	Signal CV%	Conc. Values	Conc. MeanValue	Conc. Std.Dev.	Conc. CV%
1	150	64356	1492	2%	203	199	5	3%
2		43484	2396	6%	134	128	8	6%
3		55193	2100	4%	173	168	7	4%
4		58476	3312	6%	171	179	11	6%
5		57275	6779	12%	159	175	23	13%
6		49718	1222	2%	152	149	4	3%
7		44512	2133	5%	127	132	7	6%
8		52312	2114	4%	163	158	7	5%
10		47881	8825	18%	122	143	30	21%
10		45165	4354	10%	124	134	15	11%
11		50414	1429	3%	148	152	5	3%
12		48907	3601	7%	138	147	12	8%
13		50168	3985	8%	161	151	14	9%
14		52205	203	0%	158	158	1	0%
15		50687	1481	3%	149	153	5	3%
16		52442	117	0%	158	159	0	0%
17		61306	659	1%	187	189	2	1%
18		55454	1082	2%	172	169	4	2%
19		58580	1598	3%	176	180	5	3%
20		59187	1147	2%	184	182	4	2%
21		59891	3077	5%	191	184	10	6%
22		60227	NA	NA	185	185	NA	NA
23		58181	4339	7%	168	178	15	8%
24		51319	432	1%	156	155	1	1%
		53639			Avg	163		
		5692			Stdev	19		
	Signal	10.6		Concentration	CV	11.9		



IGF-1 Clinical Correlation





9. Luteinizing Hormone

- Clinical range:
- Females - menstruating: Follicular phase 2.4 - 12.6 mIU/ml
Ovulation phase 14.0 - 95.6 mIU/ml Luteal phase 1.0 - 11.4 mIU/ml Postmenopause 7.7 - 58.5 mIU/ml
- Males: 1.7 - 8.6 mIU/ml
- Theranos assay range: 0.5 – 20 ng/mL (9 – 400 mIU/mL)
- Antibody screen: 22 antibodies
- Reference assay: BioQuant ELISA
- Precision: 3 reagent lots, 8-point standard curve: Average CV:11 %
- Precision: Assay midpoint (5 ng/mL) tested across 24 instruments: Average CV: 9 %



Luteinizing Hormone: Dose response



Sample	Nominal [LH] ng/ml in sample	RLU			Mean	Std.Dev.	CV%	Cal [LH] ng/ml in sample	% Recovery of Cal. Conc.
		1	2	3					
1	80	1127802	844348	1075637	1017998	125119	12	80.01	100
		1105411	872988	1081802					
2	40	609976	542767	568565	562601	37058	7	39.91	100
		562180	503364	588756					
3	20	269374	321966	314255	302778	23426	8	20.30	102
		295266	330480	285331					
4	10	144002	131394	158923	147868	10385	7	9.60	96
		150877	143866	158147					
5	5	77399	78458	93617	80341	7322	9	5.14	103
		81546	71541	79487					
6	2	36556	28421	32448	31101	3286	11	1.96	98
		30965	27203	31014					
7	1	15392	16118	17644	15866	1276	8	1.00	100
		16424	15847	13773					
8	0.5	7578	8241	10573	8432	1128	13	0.52	105
		7431	8371	8401					
9	0.1	2445	2147	2127	2300	155	7	0.14	137
		2337	2444	520					
10	0	500	538	424	496	47	9	0.02	NA
		536	523	456					

Theranos LH assay: ULOQ – 20 ng/mL; LLOQ – 0.5 ng/mL





Luteinizing Hormone: Stability data

Reaction Surface




Detection Stability





Luteinizing Hormone: Clinical sample correlation to Reference kit





10. Osteocalcin

- Theranos assay: Intact human Osteocalcin 1 -49 is used to develop the assay
- Assay range: 1.2 – 75 ng/mL
- Antibody screen summary


	1D	2D	3D	4D	5D	6D	7D	8D	9D	10D	11D	12D	13D
1C	Red	Red	Green	Red	Green	Red	Red	Red	Red	Red	Red	Red	Red
2C	Red	Grey	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red
3C	Red	Red	Grey	Red	Yellow	Yellow	Red	Red	Red	Red	Red	Red	Yellow
4C	Red	Red	Red	Grey	Red	Red	Red	Red	Red	Red	Red	Red	Red
5C	Red	Red	Red	Red	Grey	Red	Red	Red	Red	Red	Red	Red	Red
6C	Red	Red	Red	Red	Red	Grey	Red	Red	Red	Red	Red	Red	Red
7C	Red	Red	Red	Red	Red	Red	Grey	Red	Red	Red	Red	Red	Red
8C	Red	Red	Red	Red	Red	Red	Red	Grey	Red	Red	Red	Red	Red
9C	Red	Red	Red	Red	Red	Red	Red	Red	Grey	Red	Red	Red	Red
10C	Red	Red	Red	Red	Red	Red	Red	Red	Red	Grey	Red	Red	Red
11C	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Grey	Red	Red
12C	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Grey	Red
13C	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Red	Grey

Good dose response

Moderate dose response

No response

Osteocalcin: Impact of Matrix





11. Progesterone

- **Progesterone Clinical Range:**

Female (pre-ovulation), Male and Postmenopausal Female: less than 1 ng/mL

Female (mid-cycle): 5 to 20 ng/mL

Pregnancy 1st trimester: 11.2-90.0 ng/mL

Pregnancy 2nd trimester: 25.6-89.4 ng/mL

Pregnancy 3rd trimester: 48.4-42.5 ng/mL

- **Theranos Assay Range:** 100 – 0.3ng/mL

- **Theranos Preferred Reference Assay :** Progesterone ELISA Kit. Vendor: IBL America, distributed by Fitzgerald. Range: 40 – 0.3ng/mL

- **Theranos Secondary Reference Assay :** Progesterone EIA Kit. Vendor: Alpco. Range: 60 – 0.3ng/mL

Antibody Screening Data

Capture	Detection	
	1	2
1	X	X
2		
3		


Number of capture antibodies tested : 3
 Number of detection reagents tested: 2
 Total number of capture and detection pairs tested: 4

- X Expected good pair
- Red No Modulation
- Yellow Modulation but background or other problem
- Green Modulation, good candidate pair
- Grey Not tested



theranos
redefining healthcare

Progesterone Dilution Linearity




Summary of Results for Dilution Linearity for Progesterone Assay

Sample	Dilution	Calculated [Progesterone] ng/mL	% Recovery
1	None	11.71	[100]
	2X	5.86	96
	4X	2.97	89
	8X	1.85	95
	16X	1.49	119
2	None	31.67	[100]
	2X	18.29	113
	4X	8.68	103
	8X	3.66	81
	16X	2.21	86
3	None	39.77	[100]
	2X	18.06	90
	4X	8.71	86
	8X	5.13	98
	16X	2.06	75



Progesterone: Clinical Correlation with IBL America Kit

For a set of 32 samples, the average Theranos system concentration % CV was 11.3. Please note that samples at or below 0.3ng/mL are at the edge of the limit of detection for both the Theranos and IBL assays. $R^2 = 0.96$, slope = 0.93.



Theranos Confidential


Theranos Confidential

51



Progesterone: Precision for 8 Point Serum Curve in Duplicate Across Three Assay Lots

Avg Total Signal CV %	8.9%
Avg Total Concentration CV %	13.8%
Overall Average Intra Conc CV	11.4%
Overall Average Intra Signal CV	7.4%





12. Parathyroid Hormone


- Clinical range: normal: 8 – 65 pg/mL,
- Theranos assay range: 10 – 2000 pg/mL
- Antibody Screening summary

CAB	DAB																					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	
1																						
2																						
3																						
4																						
5																						
6																						
7																						
8																						
9																						
10																						
11																						
12																						
13																						
14																						
15																						
16																						
17																						
18																						
19																						
20																						
21																						

Good dose response
 Moderate dose response
 No response



Parathyroid hormone: Training set data








PTH: Whole blood and Plasma screen

Assay Buffer Standard curve					
PTH pg/mL	Average RLU	Stdev	%CV	Calculated pg/mL	%Recovery
2000	136063	10022	7	2223	111
1000	72512	14236	20	991	99
500	37780	4058	11	429	86
100	12142	1582	13	100	100
20	3618	595	16	21	106
0	2611	300	11		

Sample #	[PTH] pg/mL	
	W. blood	Plasma
1	19	25
2	23	55
3	19	21
4	15	18
5	15	24
6	20	21
7	17	32
8	18	19
9	29	37
10	23	32



PTH: Assay Optimization





13. TRACP 5b

- **Normal ranges (U/L)**

Healthy men - 2.1-11

Healthy premenopausal women- 1.8-10.3

Healthy postmenopausal women- 4.2-12.9

Hyperparathyroid premenopausal women- 8.1-26.3

Hyperparathyroid postmenopausal women- 11.7-21.8

Hyperthyroid premenopausal women- 5.5-17.2

Hyperthyroid postmenopausal women- 20.9-26.3

Chronic renal failure- 5.8-20.1

Capture	Detection antibody									
	Ab	1	2	3	4	5	6	7	8	9
1		Black	Red	Red	Green	Red	Red	Red	Red	Red
2		Red	Black	Red	Red	Red	Red	Red	Red	Red
3		Red	Red	Black	Red	Red	Red	Red	Red	Red
4		Green	Red	Red	Black	Yellow	Red	Green	Yellow	Green
5		Red	Red	Red	Red	Black	Red	Red	Red	Red
6		Red	Red	Red	Red	Red	Black	Red	Red	Red
7		Red	Red	Red	Green	Red	Red	Black	Red	Red
8		Red	Red	Red	Red	Red	Red	Red	Black	Red
9		Red	Red	Red	Green	Red	Red	Red	Red	Black

Good dose response

Moderate dose response

No response



TRACP5b: Antibody selection, cross reactivity

Analyte	Range	Units
TRACP5b	1.06 - 7	U/L
Bone Alk. Phos	0.78 - 200	U/L
Liver Alk. Phos	1.9 - 2000	U/L
Placental Alk. Phos.	3.2 - 10,000	U/L
Intestinal Alk. Phos	4.8 - 5000	U/L

