
From: Aleksandr Rabodzey
To: Vivek Khanna; Brian Grossman
Sent: 1/15/2014 8:06:29 AM
Subject: MDx and fda

I checked w nstg, and they did confirm that a device used for testing outside of a Clia lab has to be FDA cleared. This would apply to theranos if and when they decide to move the device out of the Clia lab

Not a big deal in my view, but something to keep in mind

Sent from my iPad