

To: =SMTP:arosendorff@gmail.com[]
From: Adam Rosendorff
Sent: Tue 11/4/2014 11:19:58 PM
Importance: Normal
Subject: FW: Redraw Request
Received: Tue 11/4/2014 11:19:59 PM

From: Steve Morin
Sent: Tuesday, November 04, 2014 3:20 PM
To: Adam Rosendorff
Subject: RE: Redraw Request

I agree, didn't trust the results, decided it was best to be conservative and ask for a redraw.

Thanks,

Steve

From: Adam Rosendorff
Sent: Tuesday, November 04, 2014 3:11 PM
To: Steve Morin
Cc: cs; Tina Lin; Daniel Young; Curtis Schneider
Subject: RE: Redraw Request

Steve

Would the fact that the sample is lipemic account for the discrepant results? I think it would merely introduce a consistent bias.

This makes me worry about the reproducibility of the assay.

Adam

From: Steve Morin
Sent: Tuesday, November 04, 2014 2:56 PM
To: Adam Rosendorff
Subject: Redraw Request

For patient for Vit D on CTN, sample integrity, lipemic, results not consistent. First run 132 next run OORH (>200).

Thanks,

Steve