


theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		


Validation of Modified Siemens Aspartate Transaminase (AST) Assay

Author(s):


Signature: 	Date: 4/11/14
Name: Paul Patel, Ph.D.	Title: Team Lead, General Chemistry


Reviewer(s):

Signature:	Date:
Name:	Title:

Signature:  SURAJ SAXENA	Date: 4/11/14
Name: Daniel Young, Ph.D.	Title: Vice President

Approver(s):

Signature: 	Date: 4/11/2014
Name: Adam Rosendorff, M.D.	Title: Laboratory Director


Sunil S. Dhawan M.D.

9/19/15

Theranos Confidential

Page 1 of 15

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

TMP-00009 Rev. A, Released 08/01/13

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

Aspartate Transaminase (AST) Plasma Assay

- I. Overview**
- II. Method Principle**
- III. Definitions and Abbreviations**
- IV. Pre-clinical Validation**
 - a. Analytical Measurement Range
 - i. Limits of Blank, Detection and Quantitation
 - ii. Linearity
 - b. Analytical Specificity
 - c. Precision
- V. Clinical Validation**
 - a. Method Comparison with Predicate
 - b. Transference and Verification of Reference Interval (Venous)
 - c. Verification of Reference Interval with Finger Stick Samples
- VI. Stability**
 - a. Reagent
 - b. Sample
 - c. Calibrators

Theranos Internal Only

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

TMP-00009 Rev. A, Released 08/01/13

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider		Date: 09/24/2013	


I. Overview

Aspartate transaminase (AST) is found in high concentrations in the liver, heart, skeletal muscle and kidneys. AST is present in both the cytoplasm and mitochondria of cells. In cases involving mild tissue injury, the predominant form of AST is the cytoplasmic form. Severe tissue damage results in increased release of the mitochondrial enzyme. High levels of AST can be found in cases such as myocardial infarction, acute liver cell damage, viral hepatitis, and carbon tetrachloride poisoning. Slight to moderate elevation of AST is seen in muscular dystrophy, dermatomyositis, acute pancreatitis and crushed muscle injuries.

II. Method Principle

The concentration of NADH is measured by its absorbance at 340/410 nm, and the rate of absorbance decrease is proportional to the AST activity. The reaction is initiated by the addition of α -ketoglutarate as a second reagent.

Reaction Equation



III. Definitions and Abbreviations

The following definitions and abbreviations are used in this document and related documents and attachments:

- a. **Accuracy:** Accuracy is defined by CLSI as the closeness of agreement between a test result and an accepted reference value. Method accuracy is used in a different sense by the American Association of Pharmaceutical Scientists where it is expressed as percent relative error (%RE). Trueness, a related CLSI term, is the closeness of agreement between the average of a number of replicate measured quantity values and a reference quantity value.
- b. **Analyte:** Component represented in the name of a measurable quantity. The closely related term measurand is defined as the particular quantity subject to measurement.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

- c. **Analytical sensitivity:** There are several alternative uses of this term. Most commonly, and for the purposes of this Validation Plan, it is used interchangeably with limit of detection. It is also used to describe the ability of an analytical method to assess small variations of the concentration of an analyte, such as the slope of the calibration curve (IUPAC).
- d. **Analytical specificity:** Ability of a test or procedure to correctly identify or quantify an entity, including in the presence of interfering substance(s) or phenomena.
- e. **Calibration:** Set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system, or values represented by a material measure or a reference material, and the corresponding values realized by standards. Under CLIA, calibration refers to the process of testing and adjusting an instrument, kit, or test system, to provide a known relationship between the measurement response and the value of the substance being measured by the test procedure (42 CFR 493.1217).
- f. **Calibrator:** A substance, material, or article intended to be used to establish the measurement relationships of a diagnostic medical device.
- g. **CLIA:** Clinical Laboratory Improvement Amendments of 1988. Congressional legislation that defined and requires specific quality assurance practices in clinical laboratories.
- h. **CLSI:** Clinical and Laboratory Standards Institute.
- i. **Coefficient of Variation:** The ratio of the standard deviation to the average, often multiplied by 100 and expressed as a percentage, abbreviated as %CV .
- j. **Colorimetry:** A technique used to determine the concentration of colored compound(s) in solution.
- k. **Interfering substance:** A substance or quantity thereof that is not the measurand but that affects the result of the measurement.
- l. **IUPAC:** International Union of Pure and Applied Chemistry
- m. **LDT:** Laboratory –developed Test.

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider		Date: 09/24/2013	

- n. **Linearity:** Linearity is the ability of a quantitative analytical method to provide results that are directly proportional to the concentrations of an analyte in test samples, within a given measuring interval. It is an important parameter to confirm when evaluating an analytical method because it verifies correct interpolation of results between points.
- o. **LMR:** Lower end of the measuring range is the lowest level at which defined conditions, including all stated characteristic of the method, are met.
- p. **LoB:** Limit of Blank is the highest value in a series of results on a sample that contains no analyte.
- q. **LoD:** Limit of Detection is the lowest amount of analyte in a sample that can be detected with stated probability, although perhaps not quantified as an exact value.
- r. **LoQ:** When used without a prefix, the Limit of Quantitation is the lowest actual concentration at which an analyte is reliably detected and at which uncertainty of the test result is less than or equal to the goal set by the manufacturer or laboratory. The term may also be used with prefixes L for lower (LLOQ) and U for upper (ULOQ), respectively. Note: $LoB < LoD \leq LoQ$.
- s. **Matrix:** All components of a material system, except the analyte. A specimen matrix is the biological milieu in which an analyte exists (e.g., plasma, serum, urine, or other body fluids).
- t. **Measuring Interval (reportable range; analytical measurement range or AMR):** A measuring interval consists of all numeric values between the lower and upper numeric values for which a method can produce quantitative results suitable for clinical use. Where applicable, a linearity study is frequently used to establish or verify the measuring interval that can be reported for a measurement method. Alternatively, the lower limit of the measuring interval may be assigned as the LoQ (LLOQ).
- u. **Precision:** Precision is the closeness of agreement between indications or measured quantity values obtained by replicate measurements on the same or similar objects under specified conditions. It is usually expressed numerically in terms of standard deviation (SD) or percent Coefficient of Variation (%CV).

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider		Date: 09/24/2013	

- v. **Reference interval:** The interval between and including two reference limits. It is common practice to define a reference limit so a stated fraction of the reference values is less than or equal, or greater than or equal, to the respective upper or lower limit.
- w. **SOP:** Standard Operating Procedure.
- x. **Spectrophotometry:** The quantitative measurement of the transmission (or reflection) properties of a material as a function of wavelength.
- y. **Testing System:** The entirety of the testing process, including instrument, sample, reagents, supplies, and procedures. Personnel are sometimes included in the definition.

IV. Pre-clinical Validation

a. Analytical Measurement Range

i. Limits of Blank, Detection and Quantitation

The limits of blank, detection and quantitation were determined to be 0.0 U/L, 6.0 U/L and 6.0 U/L respectively.

Limit of blank

CLSI guideline EP17-A section 4.3.1

Level	Number of samples	N	Mean	SD
Blank	1	20	0.0	0.0
Alpha	5%			
Parametric LoB	0.0			

Limit of detection

CLSI guideline EP17-A section 4.3.2

Level	Number of samples	N	Pooled SD
Low	1	20	3.6
Beta	5%		
Parametric LoD	6.0		

Theranos Confidential

Page 6 of 15

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

TMP-00009 Rev. A, Released 08/01/13

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

Limit of quantitation

CLSI guideline EP17-A section 5.1


Level	Number of samples	N
Low	1	20
Bias	0.9	
Pooled imprecision	3.6	
95% total error	7.9	
Allowable error	20	
LoQ	6.0	

95% total error is less than allowable error: 20 U/L.

LoQ has been established.

ii. Linearity

The Analytical Measurement Range (AMR) including linear measurement interval has been determined for Aspartate Transaminase (AST) in plasma. This method is linear from 11.6 – 1143.0 U/L within the 10% allowable non-linearity in this interval.



Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

TMP-00009 Rev. A, Released 08/01/13

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

A linear relationship fits the data better than a nonlinear relationship over the measuring interval.

Level	Mean	Linear fit	Nonlinear fit	Nonlinearity	Allowable nonlinearity
1	11.6	7.3	-	-	0.8
2	288.0	295.0	-	-	28.4
3	577.0	577.0	-	-	55.4
4	860.0	856.4	-	-	82.2
5	1143.2	1144.1	-	-	109.8

A linear relationship fits the data better than a nonlinear relationship over the measuring interval.

Nonlinearity is less than allowable nonlinearity: 10%.
Performance requirement verified over the measuring interval.

b. Analytical Specificity

The analytical specificity for this assay was determined by testing the effect of hemoglobin (100 mg/dL), bilirubin (11 mg/dL) and triglycerides (479 mg/dL) on plasma samples spiked with the interferents and then compared with un-spiked controls. AST concentration at which the interference testing was performed was at 102 U/L. Non-interference was defined as the mean result from testing of spiked samples within 10% of the mean of the un-spiked samples. Recoveries were within 94% to 104% (see table below).

Table 1. Interference Testing For AST

Analyte (U/L)	% Recovery		
	Interferent		
	Bilirubin (11 mg/dL)	Hemoglobin (100 mg/dL)	Triglycerides (479 mg/dL)
AST (102)	97	104	94

No significant interference was observed.

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description:	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider		Date: 09/24/2013	

Precision

Level = L1

Number of observations	78
Number of runs	39
Number of runs excluded	1
Number of days	20
% of days with 1 run	5%
Runs per day	2
Replicates per run	2

L1 RPT

CLSI guideline EP05-A2 section 10.4 recommends a minimum of 40 runs,

Mean	40.9		
	SD	95% CI	CV
Repeatability	2.6	2.1 to 3.3	6.2%
Between-run	0.0		0.0%
Between-day	1.9		4.7%
Within-laboratory	3.2	2.7 to 3.9	7.8%

Level = L2

Number of observations	80
Number of runs	40
Number of days	20
Runs per day	2
Replicates per run	2

Mean	111.6		
	SD	95% CI	CV
Repeatability	17.2	14.1 to 22.0	15.4%
Between-run	0.0		0.0%
Between-day	2.9		2.6%
Within-laboratory	17.4	15.1 to 20.6	15.6%

Where CV's are reported as zeros in the precision summary output this is most likely a consequence of rounding values in StatisPro.

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

Level = L3

Number of observations	76
Number of runs	38
Number of runs excluded	2
Number of days	19
Runs per day	2
Replicates per run	2

CLSI guideline EP05-A2 section 10.4 recommends a minimum of 40 runs.

Mean	256.1		
	SD	95% CI	CV
Repeatability	4.4	3.6 to 5.7	1.7%
Between-run	0.0		0.0%
Between-day	2.5		1.0%
Within-laboratory	5.1	4.3 to 6.2	2.0%

V. Clinical Validation

a. Method Comparison with Predicate (Accuracy/Comparability)

To test the accuracy of the assay on the Theranos System, 40 unique patient samples were screened on the predicate method (Siemens, Advia) and on the Theranos method. One sample was excluded as an outlier (mean absolute difference greater than 4). Using the predicate method twenty eight (28) values were within the reference range (0—34 U/L) and 11 were above the reference range. Based on the results of the data examination, either a simple linear regression or alternative procedures were used to estimate expected (average) bias and the confidence interval of expected bias at the desired medical decision level(s) as per CLSI guidance EP09-A2. StatisPro was used for bias calculations. These estimates were compared with internal criteria to judge the acceptability of the Theranos method. Each sample was run in duplicate on the predicate, and the average used for comparison to the Theranos method. Some samples were stored before analysis on both methods. If the confidence interval for the predicted bias includes the defined acceptable bias or if the acceptable bias is greater than the higher limit of the confidence interval of the predicted bias, then the data do not show that the bias of the Theranos method is different from the acceptable bias or there is a high probability (97%) that the

Theranos Confidential

Page 10 of 15

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

TMP-00009 Rev. A, Released 08/01/13

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

predicated bias is acceptable, respectively. The acceptable bias at each medical decision level was determined based on the total allowable error (TEa) minus the measured precision at the level closest to that decision level. Total allowable error (TEa) was taken from American Proficiency Institute (API) peer proficiency testing criteria or CLIA proficiency testing criteria for acceptable analytical performance, as printed in the Federal Register February 28, 1992;57(40):7002-186, when available. The TEa for AST is 20%. The table below shows the allowable bias and precision at 2 levels (values shown in parentheses) and the corresponding closest medical decision limits.

Table 2. Allowable Bias and Precision at the Medical Decision Levels

Medical Decision Levels (U/L)	20 (40.9)	60.0 (111.6)	300.0 (256.1)
Precision (%)	4.7	2.6	1.0
Allowable Bias (%)	15.3	17.4	19.0

V Method Comparison, r² = 1.00




Figure 1. Graph showing Theranos method versus Predicate Method (Siemens Advia). Simple linear regression was used to establish a slope, intercept and an r². The slope, intercept and clinical correlation were determined to be 1.08, -2.30 and 1.00 respectively.

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

Comparability

CLSI guideline EP09-A2-IR section 7

Level ID	Value	Difference	SE	95% CI	Allowable difference
	20	-0.7	1.11	-2.9 to 1.6	2.5
	60	2.6	1.08	0.4 to 4.8	7.5
	300	22.0	1.30	19.4 to 24.6	37.7

Difference is less than allowable bias: 12.55%.

The difference between the two methods is not greater than the allowable difference. The performance requirement is verified.

b. Transference and Verification of Reference Interval (Venous)

Reference ranges were modified by applying the regression equation to the lower and upper reference limits of existing reference interval to generate a new reference range. New reference ranges were verified with venous samples using fifty one (51) new normal subjects. For a reference range to pass verification, 95% of values should fall within the upper and lower reference limits and 5% or fewer values fall outside of the upper and lower reference limits. For venous verification 48 (94.1%) values fell within the new reference range and 3 (5.9%) values fell outside the new reference range. See graphs below for venous samples verification.




Figure 2. Graph showing venous sample reference range verification.

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

c. Venous versus Finger Stick Comparison

A comparison of matched venous versus finger stick samples from one hundred and ninety-two (healthy) subjects showed a positive bias with the finger stick samples.




Figure 3. Plot of matched venous versus finger stick AST results A) without bias correction and B) with bias correction.

To normalize finger stick to venous a correction factor needs to be applied to finger stick results. The following correction is used:

$$\text{Corrected Finger Stick value} = (\text{Finger stick result} - \text{Intercept}) / \text{Slope}$$

(where, slope = 1.011, intercept = 6.468)

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

d. Verification of Reference Interval with Finger Stick Samples




Figure 4. Graph showing Finger stick sample reference range verification.

New reference ranges were also verified with venous matched finger sticks from fifty three (53) new normal subjects. For finger stick verification 53 (100%) values fell within the new reference range and 0 (0%) values fell outside the new reference range. See graphs below for finger stick samples verification.

Theranos.Analyte	Anti-coagulant	Existing Reference Range (RR) (U/L)		New Reference Range (U/L)	
		Reference Range (low)	Reference Range (High)	Transferred RR (low)	Transferred RR (high)
Aspartate Aminotransferase	Lithium Heparin	0	34	6	35

The limit of quantitation for this assay is 6 U/L, therefore the new reference range determined for finger stick Aspartate aminotransferase is 6.0 – 35 U/L.

VI. Stability

a. Reagents

On-board Reagent Stability

System	Stability
ADVIA 1200	60 days
ADVIA 1650/1800	60 days
ADVIA 2400	60 days

theranos	LDT Validation Report	Theranos AST Assay	Rev:
		CL RPT-14051	1
Description	Validation Report for Modified Siemens Assay of Aspartate Transaminase (AST) in Lithium Heparin Plasma		
Originator: Curtis Schneider	Date: 09/24/2013		

For all systems, unopened reagents are stable until the expiration date printed on the product label when stored at 2° - 8°C. Do not freeze reagents.

For complete details, refer to the Methods Introduction section of the system-specific Operator's Guide.

b. Sample

Plasma samples for AST analysis are stable for 2 weeks at 2-8 °C, or at least 90 days at -20 °C.

c. Calibrators

Siemens Chemistry Calibrators should be stored at 2-8 °C, protected from light, and are stable until the expiration date on the vial label. Opened calibrators are stable for 48 hours, except for total and direct, which are stable for 8 hours.

REVISION HISTORY			
Revision Level	Effective Date	Initiator	ECO Number
A	11/06/2013	A. Rosendorf	CL ECO-00117
Section Number	Description and Justification of Changes		
All	Initial Release		

Any retransmissions, reproductions, dissemination or other use of these materials by persons or entities other than the intended recipient is prohibited. This document supersedes all earlier or previous documents unless approved in writing.